

Agilent 6890 Gas Chromatograph

Maintaining Your GC



Agilent Technologies

Notices

© Agilent Technologies, Inc. 2007

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

G1530-90010

Edition

First edition, March 2007

Printed in USA

Agilent Technologies, Inc.
2850 Centerville Road
Wilmington, DE 19808-1610 USA

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

Contents

1 About Maintaining the GC

Overview of Maintenance	10
Tools and Materials Required for Maintenance	12
Safety Information	14
Preparing the GC for Maintenance	15

2 Maintaining Capillary Columns

Consumables and Parts for Columns	18
To Install a Capillary Column Hanger	19
To Condition a Capillary Column	20
To Cut a Loop from a Column	23
To Reverse a Column and Bakeout Contaminants	24
To Attach a Capillary Column Using SilTite Metal Fittings	26
To Disconnect Fused Silica Tubing From a SilTite Fitting	29

3 Maintaining the Split/Splitless Inlet

Consumables and Parts for the Split/Splitless Inlet	32
Exploded Parts View of the Split/Splitless Inlet	35
To Install a Capillary Column with the Split/Splitless Inlet	36
To Change the Septum on the Split/Splitless Inlet	40
To Clean the Septum Seat in the Insert Assembly of the Split/Splitless Inlet	42
To Change the Liner and O-Ring on the Split/Splitless Inlet	44
To Replace the Gold Seal on the Split/Splitless Inlet	46
To Replace the Filter in the Split Vent Line	48
To Clean the Split/Splitless Inlet	49
To Bakeout Contaminants from the Split/Splitless Inlet	51

4 Maintaining the Purged Packed Inlet

Consumables and Parts for the Purged Packed Inlet	54
Exploded Parts View of the Purged Packed Inlet	57
To Install a Capillary Column with the Purged Packed Inlet	58
To Change the Septum on the Purged Packed Inlet	62

To Clean the Septum Seat in the Purged Packed Inlet	64
To Install an Adapter on the Purged Packed Inlet	66
To Change the O-Ring on the Purged Packed Inlet	68
To Change the Glass Liner on the Purged Packed Inlet	69
To Install an Insulation Cup on the Purged Packed Inlet	71
To Clean the Purged Packed Inlet	72
To Bakeout Contaminants from the Purged Packed Inlet	74
To Install a Packed Metal Column	75
To Install a Packed Column Adapter on a Detector Fitting	77
To Install a Packed Glass Column	79
To Condition a Packed Column	82
To Install Ferrules on a Packed Metal Column	84

5 Maintaining the COC Inlet

Consumables and Parts for the COC Inlet	86
Exploded Parts View of the COC Inlet	89
To Install a Capillary Column with the COC Inlet	90
To Check the Needle-to-Column Size on the COC Inlet	93
To Change a Septum on the COC Inlet	94
To Install an Insert on the COC Inlet	96
To Clean the COC Inlet	98
To Replace the Needle Support Assembly in a 7683B Injector	100
To Replace a Needle in a Syringe	103
To Replace the Fused Silica Needle in a Syringe for the COC Inlet	104
To Bakeout Contaminants from the COC Inlet	106

6 Maintaining the PTV Inlet

Consumables and Parts for the PTV Inlet	108
Exploded Parts View of the PTV Inlet	110
To Install a Capillary Column with the PTV Inlet	111
To Clean the Septumless Head on the PTV Inlet	114
To Replace the Septumless Head Teflon Ferrule on the PTV Inlet	117
To Change the Septum on the PTV Inlet	119

To Clean the Septum Seat in the Septum Head Assembly of the PTV Inlet	121
To Change the Liner on the PTV Inlet	123
To Replace the Inlet Adapter for the PTV Inlet	126
To Replace the Filter in the Split Vent Line	128
To Bakeout Contaminants from the PTV Inlet	129

7 Maintaining the VI

Consumables and Parts for the VI	132
Exploded Parts View of the VI	134
To Install a Capillary Column with the VI	135
To Attach a Sample Transfer Line to the VI	138
To Remove the VI Interface	139
To Clean the VI	140
To Install the VI Interface	141
To Replace the Filter in the Split Vent Line	142
To Bakeout Contaminants from the VI Inlet	143

8 Maintaining the FID

Consumables and Parts for the FID	146
Exploded Parts Views of the FID	149
Selecting an FID jet	151
To Attach a Capillary Column Adapter on an Adaptable FID	153
To Install a Capillary Column in the FID	155
To Replace the FID Collector Assembly	158
To Replace an FID Jet	160
To Perform Maintenance on the FID Collector Assembly	163
To Check the FID Leakage Current	171
To Check the FID Baseline	172
To Install the FID Insulation Cup Assembly (Adaptable FID Only)	173
To Install the Optional FID PTFE Chimney Insert	175
To Bakeout the FID	176

9 Maintaining the TCD

Consumables and Parts for the TCD	180
-----------------------------------	-----

To Install a Capillary Column in the TCD	182
To Install the Optional TCD Capillary Column Adapter	184
To Install a Capillary Column with the Optional TCD Capillary Column Adapter	185
To Bakeout Contaminants from the TCD	187

10 Maintaining the uECD

Important Safety Information About the uECD	190
Consumables and Parts for the uECD	192
Exploded Parts View of the uECD	194
To Replace the uECD Fused Silica Indented Mixing Liner and Install the Makeup Gas Adapter	195
To Install a Capillary Column in the uECD	198
To Install the Detector Insulating Cup	200
To Bakeout the uECD	202

11 Maintaining the NPD


Consumables and Parts for the NPD	206
Exploded Parts View of the NPD	209
Selecting an NPD jet	210
To Attach a Capillary Column Adapter on an Adaptable NPD	212
To Install a Capillary Column in the NPD	214
To Replace the NPD Bead Assembly	217
To Maintain the NPD Collector, Ceramic Insulators, and Jet	222
To Check the NPD Leakage Current	228

12 Maintaining the FPD

Consumables and Parts for the FPD	230
Exploded Parts View of the FPD	232
To Install a Capillary Column Adapter in the FPD	233
To Attach a Capillary Column to the FPD	235
To Change the FPD Wavelength Filter	237
To Remove the FPD Vent Tube	240
To Replace the FPD Ignitor	242
To Install the FPD Vent Tube and Cover	244

13 Maintaining a Valve

Consumables and Parts for Valves	246
Exploded Parts View of GC Rotary Valves	247
To Replace a Gas Sampling Valve Loop	248
To Align a Rotary Valve Rotor	250
To Replace a Rotary Valve in the Valve Box	251
To Remove the Upper Valve Box	254
To Install the Upper Valve Box	256



1

About Maintaining the GC

Overview of Maintenance	10
Tools and Materials Required for Maintenance	12
Safety Information	14
Preparing the GC for Maintenance	15

This section provides an overview of the maintenance procedures included in this document. It also lists the tools needed for routine maintenance and the safety information one should be aware of before performing a maintenance task.



Overview of Maintenance

This manual details the routine tasks needed to maintain the 6890 Gas Chromatograph (GC). The procedures assume a basic knowledge of tool use and of GC operation. Readers are, for example, expected to know how to:

- Safely turn devices on and off
- Load methods
- Change component temperatures, flows, and pressures
- Make typical pneumatic connections using Swagelok and other standard fittings

Where to find a procedure

Included in this manual are chapters on maintaining the following GC components:

- Capillary Columns
- Split/Splitless Inlet
- Purged Packed Inlet
- COC Inlet
- PTV Inlet
- Volatiles Inlet (VI)
- FID
- TCD
- uECD
- NPD
- FPD
- Valves

Each chapter includes:

- A list of the most commonly used consumables and parts for the component
- An exploded parts view of the component
- Detailed procedures for routine maintenance tasks associated with the component

Early Maintenance Feedback feature

The Agilent Lab Monitor & Diagnostic Software includes the capability to alert users of upcoming maintenance needs. This feature, called Early Maintenance Feedback, notifies users when a counter (such as a septum counter, jet cleaning counter, injection counter, or uECD wipe test counter) has reached the specified maintenance point. After performing the required maintenance, reset the applicable counter to resume using the Early Maintenance Feedback feature. Refer to the features provided by the Agilent Lab Monitor & Diagnostic Software for more details on this.

Tools and Materials Required for Maintenance

Table 1 lists the tools needed for most GC maintenance procedures. The specific tools required to perform a maintenance procedure are listed in step 1 of the procedure.

Table 1 Tools and materials for GC maintenance

Common tools
Wrench, angled, septum nut (19251-00100)*
Wrench, open-end, 1/4-inch and 5/16-inch (8710-0510)*
Wrench, open-end, 9/16-inch and 7/16-inch (8710-0803)*
Wrench, capillary inlet (G3452-20512)*
Flathead screwdriver
Column cutter, wafer (5181-8836, 4/pk)*
Driver, nut, 1/4-inch (8710-1561)*
T-20 Torx key (8710-1807) or screwdriver*
T-10 Torx key (8710-2140) or screwdriver*
3-mm hex key wrench (8710-2411)
Electronic flow meter(s) or bubble meter(s) capable of calibrated measurements at 1, 10, and 100 mL/min flow ranges.
Electronic leak detector
Magnifying loupe, 20X (430-1020)
Metric ruler
Bench vise (for setting Swagelok fittings)
Razor or sharp knife
Tweezers (8710-0007) or thin needle-nose pliers (8710-0004)
Needle-nose pliers
ESD wrist strap (for installing new components)
Gloves, heat-resistant (for handling hot parts)
Wooden cotton swab (for removing FID filters)
Tools and materials for cleaning procedures
Cleaning brushes—The FID cleaning kit (9301-0985) contains appropriate brushes for cleaning detectors and inlets
Cleaning brushes—(8710-1346) For cleaning split/splitless inlet split vent fitting, FID and collectors

Table 1 Tools and materials for GC maintenance (continued)

Jet cleaning wire (.010 inch)

Clean, lint-free cloth (to protect contamination-sensitive detector parts)

Small ultrasonic cleaning bath with aqueous detergent (for cleaning detector and inlet parts)

Gloves, clean, lint-free, nylon (large: 8650-0030, small: 8650-0029) (for handling contamination-sensitive parts)

Steel wool, 0- or 00-grade (for cleaning an inlet's septum seating surfaces)

* Included with the GC ship kits

Safety Information

Before performing a maintenance task, read the important safety and regulatory information found in the 6890 User Information book.

Preparing the GC for Maintenance

Before most maintenance procedures, the GC must be made ready. The purpose of this preparation is to avoid damage to both the instrument (electronics, columns, etc.) and the user (shocks, burns).

Column and oven preparation

The main hazards here are temperature (burns) and column exposure to air.

- Cool the oven by changing its setpoint to 35 °C. This allows the oven fan to assist cooling.
- Leave the carrier gas flow **On** until the oven has cooled. This protects the column from oxygen damage.

Inlet preparation

We are concerned with the possibility of burns and air intrusion into the column.

- *After the oven and columns have cooled*, reduce all inlet flows to 0.0 and turn the temperatures **Off**.
- For inlet-only maintenance, leave all detectors at their normal setpoints except for the TCD filament, which should be turned **Off**.
- If the column is to be removed, cap both ends to keep air out.


Detector preparation

This is another burn hazard area, plus the possibility of damage to the very sensitive electronics.

Some detectors (uECD, FPD, NPD) require 12 hours or longer to stabilize from the detector-off condition.

- To cool the detector, reduce the temperature setpoint to 35 °C.
- Some detectors (FID, NPD, FPD) use high voltages. The high voltage supply is part of the electrometer. Turn it **Off** to disable the high voltage.
- The filament in the TCD will be damaged if exposed to air while hot. To protect the filament, turn it **Off**.

1 About Maintaining the GC



2

Maintaining Capillary Columns

Consumables and Parts for Columns	18
To Install a Capillary Column Hanger	19
To Condition a Capillary Column	20
To Cut a Loop from a Column	23
To Reverse a Column and Bakeout Contaminants	24
To Attach a Capillary Column Using SilTite Metal Fittings	26
To Disconnect Fused Silica Tubing From a SilTite Fitting	29



Consumables and Parts for Columns

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 2 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

To Install a Capillary Column Hanger

WARNING


Be careful! The oven may be hot enough to cause burns. If the oven is hot, wear heat-resistant gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 1 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.
- 2 Select either the front or back hanger position. (Hanger is shown in back position.)

Front position —
Back position —



- 3 Insert the ends of the hanger into the slots in the selected position.

To Condition a Capillary Column

- 1 Gather the following:
 - One 7/16-inch, and 1/4-inch wrenches
 - No-hole ferrule (See “Consumables and Parts for Columns” on page 18.)
 - Column nut

WARNING

Do not use hydrogen as the carrier for conditioning! It could vent into the oven and present an explosion hazard.

- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING


Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Install the column into the inlet using the new ferrules. See:
 - “To Install a Capillary Column with the Split/Splitless Inlet” on page 36
 - “To Install a Capillary Column with the Purged Packed Inlet” on page 58
 - “To Install a Capillary Column with the COC Inlet” on page 90
 - “To Install a Capillary Column with the PTV Inlet” on page 111
 - “To Install a Capillary Column with the VI” on page 135

- 4 Cap the detector column fitting.



- 5 Set a minimum velocity of 30 cm/s, or as recommended by the column manufacturer. Let gas flow through the column at room temperature for 15 to 30 minutes to remove air.
- 6 Program the oven from room temperature to the maximum temperature limit for the column. Increase the temperature at a rate of 10 to 15 °C/min. Hold at the maximum temperature for 30 minutes.
- 7 Prepare the GC for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 8 Attach the column to the detector. For details, select your specific detector from the following list:
- “To Install a Capillary Column in the FID” on page 155
 - “To Install a Capillary Column in the NPD” on page 214
 - “To Install a Capillary Column in the TCD” on page 182
 - “To Install a Capillary Column in the uECD” on page 198
 - “To Install a Capillary Column Adapter in the FPD” on page 233

2 Maintaining Capillary Columns

- 9** Restore the analytical method.
 - For FID or FPD, immediately turn off the flame.
 - For NPD, immediately turn off the bead.
- 10** After the GC becomes ready, wait 10 minutes, then ignite the detector flame or bead.

To Cut a Loop from a Column

- 1 Gather the following:
 - New ferrule(s) for the column inlet connection
 - Column cutter
- 2 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Loosen the inlet column nut and remove the column from the inlet.
- 4 Uncoil one loop of column from the column hanger.
- 5 Cut the unwanted loop from the column.
- 6 Install the column into the inlet using the new ferrules. See:
 - “To Install a Capillary Column with the Split/Splitless Inlet” on page 36
 - “To Install a Capillary Column with the Purged Packed Inlet” on page 58
 - “To Install a Capillary Column with the COC Inlet” on page 90
 - “To Install a Capillary Column with the PTV Inlet” on page 111
 - “To Install a Capillary Column with the VI” on page 135

To Reverse a Column and Bakeout Contaminants

- 1 Gather the following:
 - 1/4-inch wrench
 - Column cutter
- 2 Prepare the GC for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Disconnect the column from the inlet and detector.
- 4 If necessary, cut a loop from the column. (See “To Cut a Loop from a Column” on page 23.) Do not attach the column to the inlet.
- 5 Remove the column from the hanger and reverse its position (inlet and detector ends) and place the column back on the hanger.
- 6 Attach the column to the inlet.

Select your specific inlet from the following list:

- “To Install a Capillary Column with the Split/Splitless Inlet” on page 36
- “To Install a Capillary Column with the Purged Packed Inlet” on page 58
- “To Install a Capillary Column with the COC Inlet” on page 90
- “To Install a Capillary Column with the PTV Inlet” on page 111
- “To Install a Capillary Column with the VI” on page 135

- 7 Attach your column to the detector.

Select your specific detector from the following list:

- “To Install a Capillary Column in the FID” on page 155
- “To Install a Capillary Column in the NPD” on page 214

- “To Install a Capillary Column in the TCD” on page 182
- “To Install a Capillary Column in the uECD” on page 198
- “To Install a Capillary Column Adapter in the FPD” on page 233

- 8 Set the column flow to the normal operating value, or set the capillary column gas velocity to 30 cm/s.

For split splitless, PTV, and VI inlets select split mode and set the split vent flow to 200 mL/min.

- 9 Purge the column with carrier flow for at least 10 minutes before heating the oven.
- 10 Set the inlet temperature to 300 °C or 25 °C above the normal operating temperature.
- 11 Set the column oven 25 °C above the GC method final oven temperature to bake contaminants out of the inlet, mostly through the split vent. Do not exceed the column manufacturer’s maximum temperature limit.
- 12 Bakeout for 30 minutes.

To Attach a Capillary Column Using SilTite Metal Fittings

This procedure is used to attach a capillary column to a Microfluidic splitter or switch or an Ultimate Union.

1 Gather the following:

- SilTite ferrules (see Table 3)
- Swaging nut for SilTite ferrules (G2855-20555)
- Two 1/4-inch open-end wrenches
- One 7/16-inch open-end wrench
- Column cutting tool (5181-8836)
- Internal nut (G2855-20530)
- Lint free gloves


Table 3 Available SilTite metal ferrule packages

Part number	SilTite ferrule description
5188-5361	For 0.2- to 0.25-mm columns
5188-5362	For 0.32-mm columns
5188-5363	For 0.53-mm columns


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.


- 2** Pass the tubing end through the internal nut and SilTite ferrule leaving approximately 1 cm of fused silica tubing protruding beyond the ferrule. Thread the swaging nut onto the internal nut with the tube protruding.



- 3** Using two wrenches against each other, tighten the two nuts together a little at a time, occasionally checking to see if the ferrule is gripping the tube. When the ferrule just starts to grip, notice position of the nuts and then tighten one of the nuts by turning 45 to 60 degrees of rotation, but no more than 60 degrees (one flat).



- 4** Remove the swaging nut.




2 Maintaining Capillary Columns

- 5 Using a wafer column cutter, trim the tubing at the small end of the ferrule, leaving approximately 0.3 mm of tubing extending beyond the ferrule.

Check the end of the tube with a magnifier. The end of the tube need not be perfectly square, but should not have cracks that extend under the ferrule.

NOTE


It is important that the tube end does not extend beyond 0.5 mm from the end of the ferrule.



- 6 Insert the assembled ferrule and nut into the SilTite fitting. Tighten with a wrench by only 15 to 20 degrees of rotation.


To Disconnect Fused Silica Tubing From a SilTite Fitting

Loosen and remove the internal nut. If tubing and ferrule do not come free, insert a pointed object (pen, paper clip) into the ferrule release hole and press firmly. You will hear a click as the ferrule releases.



The SilTite ferrule seal should remain leak-free for many disconnections and reconnections.

2 Maintaining Capillary Columns



3

Maintaining the Split/Splitless Inlet

Consumables and Parts for the Split/Splitless Inlet	32
Exploded Parts View of the Split/Splitless Inlet	35
To Install a Capillary Column with the Split/Splitless Inlet	36
To Change the Septum on the Split/Splitless Inlet	40
To Clean the Septum Seat in the Insert Assembly of the Split/Splitless Inlet	42
To Change the Liner and O-Ring on the Split/Splitless Inlet	44
To Replace the Gold Seal on the Split/Splitless Inlet	46
Check for leaks.	47
To Replace the Filter in the Split Vent Line	48
To Clean the Split/Splitless Inlet	49
To Bakeout Contaminants from the Split/Splitless Inlet	51



Consumables and Parts for the Split/Splitless Inlet

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 4 Split, splitless, direct, and direct connect inlet liners

Mode	Description	Deactivated	Part number
Split	Low-pressure drop, glass wool, single taper, 870 μ L	Yes	5183-4647
Split	Glass wool, 990 μ L	No	19251-60540
Split—Manual only	Empty pin and cup, 800 μ L	No	18740-80190
Split—Manual only	Packed pin and cup, 800 μ L	No	18740-60840
Splitless	Single taper, glass wool, 900 μ L	Yes	5062-3587
Splitless	Single taper, no glass wool, 900 μ L	Yes	5181-3316
Splitless	Dual taper, no glass wool, 800 μ L	Yes	5181-3315
Splitless—Direct inject	2-mm id, quartz, 250 μ L	No	18740-80220
Splitless—Direct inject	2-mm id, 250 μ L	Yes	5181-8818
Direct inject —Headspace or purge and trap	1.5-mm id, 140 μ L	No	18740-80200
Direct column connect	Single taper, splitless 4-mm id	Yes	G1544-80730
Direct column connect	Dual taper, splitless 4-mm id	Yes	G1544-80700

Table 5 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292

Table 5 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Table 6 Other consumables and parts for the split/splitless inlet


Description/quantity	Part number
Septum retainer nut for headspace	18740-60830
Septum retainer nut	18740-60835
11-mm septum, high-temperature, low-bleed, 50/pk	5183-4757
11-mm septum, prepierced, long life, 50/pk	5183-4761
Merlin Microseal septum (high-pressure)	5182-3444
Merlin Microseal septum (30 psi)	5181-8815
Nonstick fluorocarbon liner O-ring (for temperatures up to 350 °C), 10/pk	5188-5365
Nonstick fluorocarbon liner O-ring for Flip Top Inlet Sealing System, 10/pk	5188-5366
Graphite O-ring for split liner (for temperatures above 350 °C), 10/pk	5180-4168

3 Maintaining the Split/Splitless Inlet

Table 6 Other consumables and parts for the split/splitless inlet

Description/quantity	Part number
Graphite O-ring for splitless liner (for temperatures above 350 °C), 10/pk	5180-4173
Split trap vent replacement kit, 2 filters and 4 O-rings	G1544-80530
Retaining nut	G1544-20590
Gold-plated seal (standard application)	5188-5367
Gold-plated seal with cross (high split flows) (includes SS washer)	5182-9652
Stainless steel washer (0.375-inch od), 12/pk	5061-5869
Reducing nut	18740-20800
Column nut, blanking plug	5020-8294
Capillary inlet preventative maintenance kit, split	5188-6496
Capillary inlet preventative maintenance kit, splitless	5188-6497

Exploded Parts View of the Split/Splitless Inlet



To Install a Capillary Column with the Split/Splitless Inlet

WARNING

Do not use hydrogen as the carrier for conditioning! It could vent into the oven and present an explosion hazard.

- 1 Gather the following (see “Consumables and Parts for the Split/Splitless Inlet” on page 32):
 - Column
 - Ferrule(s)
 - Column nut
 - Septum
 - Column cutter
 - Isopropanol
 - Lab tissue
 - Metric ruler
 - 1/4-inch open-end wrench
 - Lint-free gloves
- 2 Prepare the GC for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Verify that the correct glass liner is installed. (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
- 4 Place the column on the hanger with the ends pointing up and the label to the front.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 5** Place a septum, capillary column nut, and ferrule on the column.




- 6** Score the column using a glass scribing tool. The score must be square to ensure a clean break.




- 7** Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.



- 8** Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 9** Position the column so it extends 4 to 6 mm above the end of the ferrule. Slide the septum up the column to hold the column nut at this position.



- 10** Thread the column nut into the inlet but do not tighten.



- 11 Adjust the column position so that the septum contacts the bottom of the column nut. Finger-tighten the column nut until it begins to grip the column.
- 12 Tighten the column nut an additional 1/4 to 1/2 turn with a wrench so that the column cannot be pulled from the fitting with gentle pressure.
- 13 Configure the new column.
- 14 Condition the column per the manufacturer's recommendation. (See "To Condition a Capillary Column" on page 20.)
- 15 Install the column into the detector. See:
 - "To Install a Capillary Column in the FID" on page 155
 - "To Install a Capillary Column in the TCD" on page 182
 - "To Install a Capillary Column in the uECD" on page 198
 - "To Install a Capillary Column Adapter in the FPD" on page 233
 - "To Install a Capillary Column in the NPD" on page 214
- 16 After the column is installed at both inlet and detector, establish a flow of carrier gas and purge as recommended by the column manufacturer.
- 17 Restore the analytical method.
 - For FID or FPD, immediately turn off the flame.
 - For NPD, immediately turn off the bead.
- 18 After the GC becomes ready, wait 10 minutes then ignite the detector flame or bead.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- 19** Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.


To Change the Septum on the Split/Splitless Inlet

- 1 Gather the following:
 - Replacement septum. (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
 - 0- or 00-grade steel wool (optional)
 - Tweezers
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

- 3 Remove the septum retainer nut or Merlin cap.
- 4 Use tweezers to remove the septum or Merlin Microseal from the retainer nut. Do not gouge or scratch the interior of the septum head.




- 5 Firmly press the new septum or Merlin Microseal into the fitting. The metal parts side of the Merlin Microseal should face down (toward the oven).



- 6 Install the septum retainer nut or Merlin cap and finger-tighten. Tighten the septum retainer nut until the C-ring is about 1 mm above the nut.

CAUTION

Overtightening the septum nut can cause contamination.



- 7 Restore the analytical method.


To Clean the Septum Seat in the Insert Assembly of the Split/Splitless Inlet

- 1 Gather the following:
 - Replacement septum (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
 - 0- or 00-grade steel wool (optional)
 - Tweezers
 - Compressed, filtered, dry air or nitrogen
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

- 3 Unscrew the insert nut from the inlet body. Lift the septum assembly straight up and away from the inlet to avoid chipping or breaking the liner.



- 4 Remove the septum retainer nut or Merlin cap.
- 5 Use tweezers to remove the septum or Merlin Microseal from the retainer nut. (See “To Change the Septum on the Split/Splitless Inlet” on page 40.)
- 6 Scrub the residue from the retainer nut and septum holder with a small piece of rolled-up steel wool and tweezers. Do not do this over the inlet.

- 7** Use compressed air or nitrogen to blow away the pieces of steel wool and septum.
- 8** Replace the insert retainer nut, tightening it to firm finger-tightness. Do not overtighten.
- 9** Firmly press the new septum or Merlin Microseal into the fitting. (See “To Change the Septum on the Split/Splitless Inlet” on page 40.)
- 10** Replace the septum retainer nut or Merlin cap and finger-tighten. (See “To Change the Septum on the Split/Splitless Inlet” on page 40.)
- 11** Restore the analytical method.


To Change the Liner and O-Ring on the Split/Splitless Inlet

- 1 Gather the following:
 - Replacement O-ring (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
 - Replacement liner
 - Tweezers
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

- 3 Unscrew the insert nut from the inlet body. Lift the septum assembly straight up and away from the inlet to avoid chipping or breaking the liner.



- 4 Loosen the O-ring from the sealing surface with tweezers.
- 5 Grasp the liner with tweezers and pull it out.
- 6 Inspect the surface of the gold seal for graphite or rubber septum contamination. If required, replace the gold seal. (See “To Replace the Gold Seal on the Split/Splitless Inlet” on page 46.)


Clean the inlet if there is visible or suspected contamination.
(See “To Clean the Split/Splitless Inlet” on page 49.)

Clean O-ring residue from sealing surface.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 7 Slide a new O-ring onto the replacement liner.
- 8 Return the liner to the inlet, pushing it all the way in until the liner contacts the gold seal.



- 9 Replace the insert retainer nut, tightening it to firm finger-tightness. Do not overtighten.
- 10 Turn on the inlet. Allow the inlet and column to purge with carrier gas for 15 minutes before heating the inlet or the column oven.
- 11 Bakeout contaminants. (See “To Bakeout Contaminants from the Split/Splitless Inlet” on page 51.)
- 12 Restore the analytical method.
- 13 Check for leaks.

To Replace the Gold Seal on the Split/Splitless Inlet

- 1 Gather the following:
 - Replacement gold seal (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
 - Replacement washer
 - 1/4-inch wrench (for column)
 - 1/2-inch wrench
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.


- 3 Remove the inlet liner.
- 4 Remove the column from the inlet. Cap the open end of the column to prevent contamination. Remove the insulation cup around the base of the inlet.
- 5 Loosen and remove the reducing nut. Remove the washer and seal inside the reducing nut.



CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 6 Put on gloves to protect the new gold seal and washer from contamination. Put a new washer in the reducing nut and place the new gold seal on top of it (raised portion facing down).



- 7** Replace the reducing nut and tighten securely with a wrench.
- 8** Replace the inlet liner.
- 9** Install the column and the insulation cup.
- 10** Bakeout contaminants. (See "To Bakeout Contaminants from the Split/Splitless Inlet" on page 51.)
- 11** Restore the analytical method.
- 12** Check for leaks.

To Replace the Filter in the Split Vent Line

- 1 Gather the following:
 - New filter cartridge. (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
 - T-20 Torx screwdriver
- 2 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

The split vent trap may contain residual amounts of any samples or other chemicals you have injected into the GC. Follow your company’s safety procedures for handling these types of substances while replacing the trap filter cartridge.

- 3 Remove the plastic pneumatics cover (top, back of GC).
- 4 Lift the filter trap assembly from the mounting bracket and unscrew the filter trap assembly.
- 5 Remove the old filter cartridge and two O-rings.



- 6 Verify the new O-rings are seated properly on the new filter cartridge.
- 7 Install the new filter cartridge then reassemble the trap.
- 8 Place the filter trap assembly in the mounting bracket.
- 9 Fully tighten the split vent front weldment onto the trap.
- 10 Check for leaks.
- 11 Restore the analytical method.

To Clean the Split/Splitless Inlet

- 1 Gather the following:
 - Replacement septum (See “Consumables and Parts for the Split/Splitless Inlet” on page 32.)
 - Replacement liner
 - Replacement O-ring
 - Replacement gold seal
 - Replacement washer
 - Solvent that will clean the type of deposits in your inlet
 - Compressed, filtered, dry air or nitrogen
 - Beaker
 - Cleaning brushes—The FID cleaning kit (part number 9301-0985) contains appropriate brushes
 - Lint-free gloves
- 2 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

- 3 Remove the inlet liner. (See “To Change the Liner and O-Ring on the Split/Splitless Inlet” on page 44.)
- 4 Disconnect the column from the inlet.
- 5 Remove the reducing nut and gold seal. (See “To Replace the Gold Seal on the Split/Splitless Inlet” on page 46.)
- 6 Place a beaker in the oven under the inlet to catch the solvent.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 7 Soak a cleaning brush in the solvent and scrub the inside of the inlet weldment. Repeat 10 times.
- 8 Rinse the inlet with the solvent.
- 9 Blow the inside of the inlet dry with compressed air or nitrogen.

3 Maintaining the Split/Splitless Inlet

- 10** Install the gold seal and reducing nut.
- 11** Install the liner and O-ring.
- 12** Install the column. (See “To Install a Capillary Column with the Split/Splitless Inlet” on page 36.)
- 13** Check for leaks.
- 14** Bakeout contaminants. (See “To Bakeout Contaminants from the Split/Splitless Inlet” on page 51.)
- 15** Restore the analytical method.

To Bakeout Contaminants from the Split/Splitless Inlet

- 1 Put the inlet into split mode.
- 2 Set the column flow to the normal operating value, or set the capillary column gas velocity to 30 cm/s.
- 3 Set the inlet split vent flow to 200 mL/min.
- 4 Purge the column with carrier flow for at least 10 minutes before heating the oven.
- 5 If the column is attached to the detector, set the detector 25 °C above normal operating temperature.


WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

If the column is not attached to the detector, cap the detector fitting.

- 6 Set the inlet temperature to 300 °C or 25 °C above the normal operating temperature to bakeout contaminants from the inlet, mostly through the split vent.
- 7 Set the column oven 25 °C above the GC method final oven temperature to bake contaminants from the column. Do not exceed the column manufacturer's maximum temperature limit.
- 8 Bakeout for 30 minutes or until the detector baseline is free of contamination peaks.

3 Maintaining the Split/Splitless Inlet



4

Maintaining the Purged Packed Inlet

Consumables and Parts for the Purged Packed Inlet	54
Exploded Parts View of the Purged Packed Inlet	57
To Install a Capillary Column with the Purged Packed Inlet	58
To Change the Septum on the Purged Packed Inlet	62
To Clean the Septum Seat in the Purged Packed Inlet	64
To Install an Adapter on the Purged Packed Inlet	66
To Change the O-Ring on the Purged Packed Inlet	68
To Change the Glass Liner on the Purged Packed Inlet	69
To Install an Insulation Cup on the Purged Packed Inlet	71
To Clean the Purged Packed Inlet	72
To Bakeout Contaminants from the Purged Packed Inlet	74
To Install a Packed Metal Column	75
To Install a Packed Column Adapter on a Detector Fitting	77
To Install a Packed Glass Column	79
To Condition a Packed Column	82
To Install Ferrules on a Packed Metal Column	84



Consumables and Parts for the Purged Packed Inlet

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 7 Purged packed inlet parts

Description	Part number/quantity
Preventative maintenance kit	5188-6498
Purged packed glass liners and column adapters	
Glass liner	5080-8732 (25/pack) or 5181-3382 deactivated (5/pack)
0.53-mm column adapter	19244-80540
1/8-inch column adapter	19243-80530
1/4-inch column adapter	19243-80540
Recommended septa and O-rings for the purged packed inlet	
11-mm solid septum, low-bleed, red	5181-1263 (50/pk)
11-mm septum with partial through-hole, low-bleed, red	5181-3383 (50/pk)
11-mm septum, low-bleed, gray	5080-8896 (50/pk)
Merlin Microseal septum (30 psi)	5181-8815
11-mm high-temperature silicone septum (350 °C and higher)	5182-0739 (50/pk)
Viton O-ring (Top insert weldment)	5080-8898 (12/pk)

Table 8 Nuts and ferrules for packed columns

Description	Typical use	Part number/quantity
1/8-inch id Swagelok stainless steel nut, front ferrule, back ferrule	1/8-inch column	5080-8751 (20 each/pk)
1/8-inch id Swagelok brass nut, front ferrule, back ferrule	1/8-inch column	5080-8750 (20 each/pk)
1/8-inch id Vespel/ graphite ferrule	1/8-inch column	0100-1332 (10/pk)

Table 8 Nuts and ferrules for packed columns (continued)

Description	Typical use	Part number/quantity
1/8-inch id brass tubing nut	1/8-inch column	5180-4103 (10/pk)
1/4-inch id Swagelok stainless steel nut, front ferrule, back ferrule	1/4-inch column	5080-8753 (20 each/pk)
1/4-inch id Swagelok brass nut, front ferrule, back ferrule	1/4-inch column	5080-8752 (20 each/pk)
1/4-inch id Vespel/ graphite ferrule	Inlet/detector liner/adapters 1/4-inch column	5080-8774 (10/pk)
1/4-inch id brass tubing nut	1/4-inch column	5180-4105 (10/pk)

Table 9 Nuts, ferrules, and hardware for capillary columns


Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)

4 Maintaining the Purged Packed Inlet

Table 9 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Exploded Parts View of the Purged Packed Inlet



To Install a Capillary Column with the Purged Packed Inlet

- 1 Gather the following:
 - Column
 - Ferrule (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Column nut
 - Glass liner
 - Viton O-ring
 - 0.53-mm column adapter
 - Septum
 - 1/4-inch wrench
 - Metric ruler
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Install a 0.53-mm column adapter. (See “To Install an Adapter on the Purged Packed Inlet” on page 66.)


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Install a new Viton O-ring. (See “To Change the O-Ring on the Purged Packed Inlet” on page 68.)
- 5 Place a septum, capillary column nut, and ferrule on the column.




- 6** Score the column using a glass scribing tool. The score must be square to ensure a clean break.




- 7** Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.



- 8** Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 9** Position the column so it extends 1 to 2 mm above the end of the ferrule. Slide the septum up the column to hold the column nut at this fixed position.



- 10** Thread the column nut into the inlet adapter but do not tighten.



- 11** Adjust the column position so that the septum is even with the bottom of the column nut. Finger-tighten the column nut until it begins to grip the column.
- 12** Tighten the column nut an additional 1/4 to 1/2 turn with a wrench so that the column cannot be pulled from the fitting with gentle pressure.
- 13** Configure the new column.
- 14** Condition the column per the manufacturer's recommendation. (See "To Condition a Capillary Column" on page 20.)
- 15** Install the column into the detector. See:
- "To Install a Capillary Column in the FID" on page 155
 - "To Install a Capillary Column in the TCD" on page 182
 - "To Install a Capillary Column in the uECD" on page 198
 - "To Install a Capillary Column Adapter in the FPD" on page 233
 - "To Install a Capillary Column in the NPD" on page 214
- 16** After the column is installed at both inlet and detector, establish a flow of carrier gas and purge as recommended by the column manufacturer.
- 17** Restore the analytical method.
- For FPD, immediately turn off the flame.
 - For NPD, immediately set the bead voltage to 0.0.
- 18** After the GC becomes ready, wait 10 minutes then ignite the detector flame or adjust offset on the NPD bead.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- 19 Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.


To Change the Septum on the Purged Packed Inlet

- 1 Gather the following:
 - Replacement septum (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Septum nut wrench
 - 0- or 00-grade steel wool (optional)
 - Tweezers
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

- 3 Remove the septum retainer nut or Merlin cap.
- 4 Use tweezers to remove the septum or Merlin Microseal from the retainer nut. Do not gouge or scratch the interior of the septum head.




- 5 Firmly press the new septum or Merlin Microseal into the fitting. The metal parts side of the Merlin Microseal should face down (toward the oven).



- 6 Replace the septum retainer nut or Merlin cap and finger-tighten. Tighten the septum retainer nut until the C-ring is about 1 mm above the nut.

CAUTION

Overtightening the septum nut can cause contamination.



- 7 Restore the analytical method.

To Clean the Septum Seat in the Purged Packed Inlet

- 1 Gather the following:
 - Replacement septum (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Septum nut wrench
 - 0- or 00-grade steel wool (optional)
 - Tweezers
 - Compressed, filtered, dry air or nitrogen
 - Ultrasonic cleaning bath
 - Lint-free gloves
- 2 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

- 3 Remove the septum retainer nut or Merlin cap.
- 4 Loosen the top insert weldment and remove.

CAUTION


Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 5 Use tweezers to remove the septum or Merlin Microseal from the top insert weldment. Do not gouge or scratch the interior of the septum head.
- 6 Scrub the residue from the top insert weldment and septum nut with a small piece of rolled-up steel wool and tweezers. Ultrasonically clean the retainer nut and top insert weldment.
- 7 Use compressed air or nitrogen to blow away the pieces of steel wool and septum.
- 8 Wearing gloves, inspect the O-ring and replace, if necessary. (See “To Change the O-Ring on the Purged Packed Inlet” on page 68.)
- 9 Install the top insert weldment and hand-tighten firmly.

- 10 Firmly press the new septum or Merlin Microseal into the fitting.
- 11 Install the septum retainer nut or Merlin cap and finger-tighten. Tighten the septum retainer nut until the C-ring is about 1 mm above the nut.

CAUTION

Overtightening the septum nut can cause contamination.



- 12 Restore the analytical method.

To Install an Adapter on the Purged Packed Inlet

- 1** Gather the following:
 - Brass tubing nut (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Adapter (0.53 mm, 1/8-inch packed, or 1/4-inch packed)
 - 7/16-inch and 9/16-inch wrench
 - Vespel/graphite ferrule
 - Methanol
 - Lint-free gloves
- 2** Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3** Clean the end of the adapter with a lint-free cloth and methanol to remove contamination such as fingerprints.
- 4** Place the tubing nut and Vespel/graphite ferrule on the adapter.



- 5** Insert the adapter straight into the inlet base as far as possible.
- 6** Hold the adapter in this position and finger-tighten the nut.
- 7** Tighten an additional 1/4 turn with a wrench.

To Change the O-Ring on the Purged Packed Inlet

- 1 Gather the following:
 - Replacement O-ring (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Septum nut wrench
 - Tweezers
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING


Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

- 3 Loosen the top insert weldment to remove the top portion of the inlet.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Use tweezers to remove the old O-ring.



- 5 Insert a new O-ring.
- 6 Install and tighten the top insert weldment.
- 7 Restore the analytical method.


To Change the Glass Liner on the Purged Packed Inlet

- 1 Gather the following:
 - Replacement glass liner (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - 9/16-inch wrench
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

- 3 Loosen the top insert weldment to remove the top portion of the inlet.



- 4 Use a thin wire or wood splint to carefully lift and remove the old glass liner.

CAUTION


Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

4 Maintaining the Purged Packed Inlet


- 5** Wearing gloves, inspect the O-ring and replace, if necessary. (See “To Change the O-Ring on the Purged Packed Inlet” on page 68.)
- 6** Wearing gloves, grasp the flared end (top) of the replacement glass liner with tweezers and install it in the inlet. If the glass liner does not seat properly because a capillary column is installed, remove the column, install the glass liner, and replace the column. (See “To Install a Capillary Column with the Purged Packed Inlet” on page 58.)
- 7** Install the top insert weldment and hand-tighten firmly.
- 8** Restore the analytical method.

To Install an Insulation Cup on the Purged Packed Inlet


- 1 Gather the following:
 - No-hole ferrule
 - Column nut
- 2 Install a plug (for example, a column nut with no-hole ferrule) in the inlet capillary adapter.



- 3 Push the cup spring to the right. Slide the cup over the inlet fitting so that the insulation at the top of the cup is flush against the oven roof.



- 4 Place the spring into the groove in the inlet liner. Remove the column nut and no-hole ferrule.



To Clean the Purged Packed Inlet

- 1 Gather the following:
 - Replacement O-ring (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Replacement glass liner
 - Replacement septum
 - Solvent that will clean the type of deposits in your inlet
 - Compressed, filtered, dry air or nitrogen
 - Beaker
 - Cleaning brushes—The FID cleaning kit (part number 9301-0985) contains appropriate brushes
 - Lint-free gloves
- 2 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If the inlet is hot, wear heat-resistant gloves to protect your hands.

- 3 Remove the column.
- 4 Remove the septum nut and septum.
- 5 Remove the top insert weldment.
- 6 Remove the glass liner and O-ring.
- 7 If used, remove the adapter.
- 8 Ultrasonically clean the septum nut, top insert weldment, and adapter (if used) in a suitable solvent.
- 9 Place a beaker in the oven under the inlet to catch the solvent.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 10 Soak a cleaning brush with the solvent and vigorously scrub the interior walls of the inlet.
- 11 Blow the inside of the inlet dry with compressed air or nitrogen.

- 12 Install the adapter, if used. (See “To Install an Adapter on the Purged Packed Inlet” on page 66.)
- 13 Install the glass liner and O-ring. (See “To Change the Glass Liner on the Purged Packed Inlet” on page 69.)
- 14 Install the top insert weldment and finger-tighten.
- 15 Install the septum and septum nut. (See “To Change the Septum on the Purged Packed Inlet” on page 62.)
- 16 Attach the column. (See “To Install a Capillary Column with the Purged Packed Inlet” on page 58.)
- 17 Check for leaks.
- 18 Restore the analytical method.

To Bakeout Contaminants from the Purged Packed Inlet

- 1** Set the column flow to the normal operating value, or set the capillary column gas velocity to 30 cm/s.
- 2** Purge the column with carrier flow for at least 10 minutes before heating the oven.
- 3** If the column is attached to the detector, set the detector 25 °C above normal operating temperature.

If the column is not attached to the detector, cap the detector fitting.

- 4** Set the inlet temperature to 300 °C or 25 °C above the normal operating temperature.
- 5** Set the column oven 25 °C above the GC method final oven temperature to bake contaminants out of the inlet. Do not exceed the column manufacturer's maximum temperature limit.
- 6** Bakeout for 30 minutes or until the detector baseline is free of contamination peaks.

To Install a Packed Metal Column

- 1 Gather the following:
 - 7/16-inch, 9/16-inch, and 1/2-inch wrenches
 - Lint-free gloves
- 2 Prepare the GC for maintenance. See “Preparing the GC for Maintenance” on page 15.
- 3 Prepare the packed metal column. (See “To Install Ferrules on a Packed Metal Column” on page 84.)

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If the oven, inlet, or detector is hot, wear gloves to protect your hands.

- 4 Install the 1/8-inch or 1/4-inch packed column inlet adapter, if necessary. (See “To Install an Adapter on the Purged Packed Inlet” on page 66.)

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 5 Attach the column to the inlet adapter. Finger-tighten the nut.
- 6 Tighten the nut an additional 1/4 turn with a wrench (for a 1/8-inch column) or an additional 3/4 turn (for a 1/4-inch column).

Use two wrenches, one on the column nut and the other on the adapter, to prevent the adapter from rotating.

- 7 Press **[Config] [Col 1]** or **[Config] [Col 2]** and enter 0.00 in either **Length** or **Diameter**, and identify the inlet and detector to which the column is attached.

WARNING

Do not use hydrogen as the carrier for conditioning! It could vent into the oven and present an explosion hazard.

- 8 Condition the column, if necessary. (See “To Condition a Packed Column” on page 82.)

- 9 Load the GC maintenance method and wait for the GC to become ready.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If the oven, inlet, or detector is hot, wear gloves to protect your hands.

- 10 If required, install the detector adapter. (See “To Install a Packed Column Adapter on a Detector Fitting” on page 77.)
- 11 Attach the column to the detector or detector adapter. Finger-tighten the nut.
- 12 Tighten the nut an additional 1/4 turn with a wrench (for a 1/8-inch column) or an additional 3/4 turn (for a 1/4-inch column).
- 13 Establish a flow of carrier gas and purge as recommended by the packing manufacturer. Generally:
 - 20 to 30 mL/min for 2-mm id glass or 1/8-inch od metal columns
 - 50 to 60 mL/min for 4-mm id glass or 1/4-inch od metal columns
- 14 Restore the analytical method.
 - For FPD, immediately turn off the flame.
 - For NPD, immediately set the bead voltage to 0.0.
- 15 After the GC becomes ready, wait 10 minutes then ignite the detector flame or adjust offset on the NPD bead.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- 16 Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.

To Install a Packed Column Adapter on a Detector Fitting

- 1 Gather the following:
 - 7/16-inch, 9/16-inch, and 1/2-inch wrenches
 - Vespel/graphite ferrule (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Brass column nut
 - Lint-free gloves
 - Adapter. Select the appropriate adapter from one of the parts lists shown below:
 - “Consumables and Parts for the FID” on page 146 (Packed columns can only be installed on an adaptable FID.)
 - “Consumables and Parts for the TCD” on page 180
 - “Consumables and Parts for the NPD” on page 206
 - “Consumables and Parts for the FPD” on page 230
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

CAUTION


Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble a nut and a ferrule onto the adapter.



- 4 Insert the adapter straight into the detector base as far as possible.

4 Maintaining the Purged Packed Inlet



- 5** Hold the adapter in this position and finger-tighten the nut.
- 6** Tighten an additional 1/4 turn with a wrench (for a 1/8-inch column) or an additional 3/4 turn (for a 1/4-inch column).

To Install a Packed Glass Column

- 1 Gather the following:
 - 9/16-inch wrench
 - Two 1/4-inch brass nuts (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Two 1/4-inch Vespel/graphite ferrules
 - Lint-free gloves
- 2 Prepare the GC for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If the oven, inlet, or detector is hot, wear gloves to protect your hands.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble a brass nut and Vespel/graphite ferrule on each end of the column.



Glass columns must be simultaneously inserted into the inlet and detector and installed parallel to the oven door. When conditioning the column, do not attach the column to the detector.

- 4 If conditioning the column, insert the column into the purged packed inlet until it bottoms. Withdraw the column 1 to 2 mm. Finger-tighten the inlet column nut. (See “To Condition a Packed Column” on page 82.)

CAUTION


Overtightening the column nut or forcing it to bottom in either the inlet or detector may shatter the column.

- 5 Tighten the inlet column nut 1/4 turn with a wrench.

WARNING

Do not use hydrogen as the carrier for conditioning! It could vent into the oven and present an explosion hazard.

- 6 After conditioning, remove the column from the inlet.
- 7 Simultaneously insert the column into the inlet and detector fittings but *do not* force it. It may be necessary to start the long end of the column in the inlet at an angle to clear the oven floor.



- 8 Withdraw the column 1 to 2 mm from both the inlet and detector. Finger-tighten both column nuts.

CAUTION

Overtightening the column nut or forcing it to bottom in either the inlet or detector may shatter the column.

- 9 Tighten both column nuts 1/4 turn with a wrench.
- 10 Press [**Config**] [**Column #**], enter **0.00** in either **Length** or **Diameter**, and identify the inlet and detector to which the column is connected.
- 11 Establish a flow of carrier gas and purge as recommended by the packing manufacturer. Generally:
 - 20 to 30 mL/min for 2-mm id glass or 1/8-inch od metal columns
 - 50 to 60 mL/min for 4-mm id glass or 1/4-inch od metal columns
- 12 Restore the analytical method.
 - For FPD, immediately turn off the flame.
 - For NPD, immediately set the bead voltage to 0.0.

- 13 After the GC becomes ready, wait 10 minutes then ignite the detector flame or adjust offset on the NPD bead.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- 14 Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.

To Condition a Packed Column

- 1 Gather the following:
 - Capillary adapter, column nut, and no-hole ferrule (for FID and NPD), or 1/8-inch Swagelok cap (for TCD and FPD)
 - Two 7/16-inch wrenches
 - 1/4-inch open-end wrench
 - Lint-free gloves

WARNING

Do not use hydrogen as the carrier for conditioning! It could vent into the oven and present an explosion hazard.

- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If the oven, inlet, or detector is hot, wear gloves to protect your hands.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Install the proper liner in the inlet and attach the column. (See “To Install a Packed Metal Column” on page 75.)
- 4 Cap the detector(s) fittings with the capillary adapter, no-hole ferrule and column nut (FID and NPD) or 1/8-inch cap (FPD and TCD).
- 5 Enter a column flow as recommended by the packing manufacturer or an appropriate flow as follows:
 - 20 to 30 mL/min for 2-mm id glass or 1/8-inch od metal columns
 - 50 to 60 mL/min for 4-mm id glass or 1/4-inch od metal columns
- 6 Raise the oven temperature slowly to the conditioning temperature for the column. The conditioning temperature is never higher than the maximum temperature limit for the column; 30 °C less than the maximum is usually sufficient.

- 7** Continue conditioning overnight at the final temperature. Cool the oven to room temperature with carrier flow on.
- 8** Attach the column to the detector and maintain established flow. (See “To Install a Packed Metal Column” on page 75.)


To Install Ferrules on a Packed Metal Column

- 1 Gather the following:
 - Wrenches
 - Stainless steel male Swagelok fitting, 1/4- or 1/8-inch od
 - Brass Swagelok nut and ferrule set (See “Consumables and Parts for the Purged Packed Inlet” on page 54.)
 - Lint-free gloves
- 2 Verify that the column end is cut square and is free of burns and deformation.
- 3 Secure the fitting in a bench vise.


CAUTION


Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Assemble a Swagelok nut and ferrules onto the column.



- 5 Fully insert the column into the vise-held fitting, then withdraw 1–2 mm. Finger-tighten the nut.
- 6 Tighten the nut an additional 3/4 turn with a wrench (for a 1/8-inch column) or an additional 1-1/4 turn (for a 1/4-inch column).
- 7 Unscrew the column nut from the vise-held fitting and remove the column. Ferrules should now be set in place on the column with the column end correctly positioned.





5 Maintaining the COC Inlet

Consumables and Parts for the COC Inlet	86
Exploded Parts View of the COC Inlet	89
To Install a Capillary Column with the COC Inlet	90
To Check the Needle-to-Column Size on the COC Inlet	93
To Change a Septum on the COC Inlet	94
To Install an Insert on the COC Inlet	96
To Clean the COC Inlet	98
To Replace the Needle Support Assembly in a 7683B Injector	100
To Replace a Needle in a Syringe	103
To Replace the Fused Silica Needle in a Syringe for the COC Inlet	104
To Bakeout Contaminants from the COC Inlet	106



Consumables and Parts for the COC Inlet

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 10 Recommended septum nut and inserts for injections onto 0.53-mm columns

Column type	Part number
Insert, fused silica, 0.53-mm id	19245-20580 (no rings)
Insert, aluminum-clad, 0.53-mm id	19245-20780 (4 rings)
Septum nut, 530 μ m	G1545-80520
Needle support assembly, 530 μ m, for 7683B injector	G2913-60977

Table 11 Recommended parts for injections onto 0.25-mm and 0.32-mm columns

Column type	Part number
Insert, fused silica, 0.32-mm id	19245-20525 (5 rings)
Insert, 0.25-mm id	19245-20515 (6 rings)
Septum nut, 250/320 μ m	19245-80521
Syringe barrel, removable needle, 5 μ L	5182-0836
Needle, 250 μ m (3/pk)	5182-0833
Needle, 320 μ m (3/pk)	5182-0831
Needle support assembly, 250/320 μ m, for 7683B injector	G2913-60978

Table 12 Recommended parts for injections onto 0.2-mm columns

Description	Part number/quantity
Insert, fused silica, 0.20-mm id	19245-20510
Cooling tower assembly	19230-80625
Syringe barrel, for fused silica needle, 10 μ L	9301-0658
Replacement needles, fused silica, 0.18 mm	19091-63000 (6/pk)
Replacement Teflon ferrule for fused silica syringe	0100-1389
Removable stainless steel needle syringe, 10 μ L	5182-9633

Table 12 Recommended parts for injections onto 0.2-mm columns

Description	Part number/quantity
Replacement stainless steel needles, 0.23 mm	5182-9645 (3/pk)

Table 13 Recommended septa for the COC inlet

Description	Part number/quantity
For 0.53-mm and 0.25/0.32-mm septum nuts	
5-mm solid septum for manual and automatic injection	5181-1261
5-mm long-life septum	5183-4762 (50/pk)
5-mm advanced green septum	5183-4760 (50/pk)
5-mm, high-temperature, low-bleed septum	5183-4758 (50/pk)
5-mm through-hole septum for automatic injection	5181-1260 (25/pk)
For the duckbill septum	
Duckbill septum for manual injection only (must use cooling tower with duckbill)	19245-40050 (10/pk)

Table 14 Nuts, ferrules, and hardware for capillary columns


Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)

5 Maintaining the COC Inlet

Table 14 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Exploded Parts View of the COC Inlet



To Install a Capillary Column with the COC Inlet

- 1 Gather the following:
 - Column nut and ferrule. (See “Consumables and Parts for the COC Inlet” on page 86.)
 - Column cutter
 - 1/4-inch and 5/16-inch wrenches
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.


WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Before installing the column, be sure the correct insert is installed for the needle and column. (See “To Install an Insert on the COC Inlet” on page 96.)
- 4 Place a capillary column nut and ferrule on the column.




- 5 Score the column using a glass scribing tool. The score must be square to ensure a clean break.



- 6 Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.



- 7 Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 8 Gently insert the column into the inlet until it bottoms.
- 9 Insert the column nut into the inlet fitting and finger-tighten.



WARNING

To avoid bending the inlet, always use two wrenches. Use a 5/16-inch wrench to support the inlet while tightening the column nut with a 1/4-inch wrench.

- 10 Tighten the column nut an additional 1/4 turn with a wrench or until the column does not move.
- 11 If using an automatic injection system with a 0.25-mm or 0.32-mm column, verify that the column installation by manually pushing the syringe into the inlet.
- 12 Configure the new column.
- 13 Condition the column per the manufacturer's recommendation. (See "To Condition a Capillary Column" on page 20.)

- 14** Install the column into the detector. See:
 - “To Install a Capillary Column in the FID” on page 155
 - “To Install a Capillary Column in the TCD” on page 182
 - “To Install a Capillary Column in the uECD” on page 198
 - “To Install a Capillary Column Adapter in the FPD” on page 233
 - “To Install a Capillary Column in the NPD” on page 214
- 15** After the column is installed at both inlet and detector, establish a flow of carrier gas and purge as recommended by the column manufacturer.
- 16** Restore the analytical method.
 - For FPD, immediately turn off the flame.
 - For NPD, immediately set the bead voltage to 0.0.
- 17** After the GC becomes ready, wait 10 minutes then ignite the detector flame or adjust offset on the NPD bead.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- 18** Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.

To Check the Needle-to-Column Size on the COC Inlet

- 1 Gather the following:
 - Insert (See “Consumables and Parts for the COC Inlet” on page 86.)
 - Syringe needle
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Check the needle-to-column size to make certain that the needle fits in the column.
- 4 Identify the correct insert for the column size. (See “Consumables and Parts for the COC Inlet” on page 86.) Use the insert that is the same size as the syringe needle to verify that the column you plan to use is the correct size.
- 5 Insert the column into one end of the insert.



- 6 Insert the syringe needle through the other end of the insert and into the column. If the needle cannot pass easily into the column, reverse the insert to try the needle and column in the other end.

To Change a Septum on the COC Inlet

- 1 Gather the following:
 - Replacement septum. (See “Consumables and Parts for the COC Inlet” on page 86.)
 - Tweezers
 - A thin wire (0.2-inch diameter) for removing septum from inlet
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Replace the septum.
 - If you are using a septum nut, grasp the knurling and unscrew. Remove the old septum with tweezers. Use tweezers to install a new septum. Push the septum into the septum nut until properly seated. Firmly tighten the nut.


For 250/320- μ m
automated injections

For 530- μ m automated
injections



- If you are using a cooling tower, grasp the three rings and unscrew. The spring and duckbill septum may pop out of the inlet when you remove the cooling tower. Be careful not to lose them. If they do not pop out, use a thin wire to remove them from the inlet. Insert the replacement duckbill septum into the spring and place in the inlet. Reattach the cooling tower assembly, then finger-tighten.

For manual 200- μ m injections with fused silica needle



- 4 Before making an injection, check the alignment of the entire assembly using the proper size syringe.
- 5 Restore the analytical method.

To Install an Insert on the COC Inlet

- 1 Gather the following:
 - Lint-free gloves
 - Replacement insert. (See “Consumables and Parts for the COC Inlet” on page 86.)
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Remove the column from the inlet.
- 4 Locate the septum nut or cooling tower assembly at the top of the inlet and remove it. If the septum remains in the septum nut, do not remove it unless you want to change it. If necessary, replace the existing septum or duckbill with a new one. (See “To Change a Septum on the COC Inlet” on page 94.)
- 5 Remove the spring from the inlet with an extraction wire and set it aside. Be careful not to lose or damage it because you will use the spring to keep the new insert in position.
- 6 Remove the existing insert from the inlet by gently pushing it out from below with a wire or piece of column. Store the insert for possible later use.



- 7** Check that the insert is the correct size for both the needle and column. (See “To Check the Needle-to-Column Size on the COC Inlet” on page 93.)
- 8** Lower the new insert straight into the inlet from the top.
- 9** Install the spring on top of the insert.
- 10** Install the septum and septum nut or duckbill septum and cooling tower assembly and finger-tighten.
- 11** Install the column. (See “To Install a Capillary Column with the COC Inlet” on page 90.)

To Clean the COC Inlet

- 1 Gather the following:
 - 1/4-inch and 5/16-inch wrenches
 - Cleansing bath
 - Aqueous detergent
 - Distilled water
 - Methanol
 - Compressed, filtered, dry air or nitrogen
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.


WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Remove the column.
- 4 Remove the septum nut or cooling tower and then remove the septum.
- 5 Remove the existing insert from the inlet by gently pushing it out from below with a wire or piece of column. Store the insert for possible later use.



- 6 Fill an ultrasonic cleansing bath with aqueous detergent and place the spring and insert into it. Sonicate for 1 minute.
- 7 Drain the aqueous detergent and fill the bath with distilled water. Sonicate for 1 minute.
- 8 Remove the spring and insert from the bath and rinse thoroughly with water and methanol.
- 9 Dry the spring and insert with compressed air or nitrogen.
- 10 Install the insert. (See “To Install an Insert on the COC Inlet” on page 96.)
- 11 Install the column. (See “To Install a Capillary Column with the COC Inlet” on page 90.)

To Replace the Needle Support Assembly in a 7683B Injector

- 1 Gather a 7683B needle support assembly for injections onto 530- μm or 250/320- μm columns.




250 μm /320 μm
(G2913-60978)




530 μm
(standard, G2913-60978)

- 2 Remove all vials and bottles from the turret, and disconnect the injector cable from the GC.
- 3 Open the injector door.
- 4 Remove the syringe.
- 5 With your finger under the shaft near the bearing on the needle support assembly, pull gently to release the bearing from the bearing clip in the syringe carriage.

**CAUTION**

Do not to pull the assembly by its metal shaft, as it is easily bent.

- 6 Carefully use the bearing to pull the rod down until you can lift the assembly out of the syringe carriage.
- 7 To install the needle support assembly, use your right hand to insert the upper end of the rod into the plastic guide to the right of the plunger carrier loop.
- 8 Turn the needle support assembly so that the flat surface of the slide glides up and down the syringe carriage tracks.



- 9** Align the bearing on the needle support assembly with the plastic bearing clip to the right of the syringe latch and push gently on the bearing until the assembly snaps into place.

CAUTION


Be careful not to bend the needle during installation.

Do not operate the injector without a syringe or align the probe in place because the free-swinging syringe latch may interfere with the motor and jam the syringe carriage.

-
- 10** Install the syringe.

To Replace a Needle in a Syringe

- 1 Gather the following (see Table 11, “Recommended parts for injections onto 0.25-mm and 0.32-mm columns,” on page 86):
 - Syringe barrel
 - Needle, 250- μ m or 320- μ m
- 2 Unscrew the syringe barrel cap and remove the spring.
- 3 Make sure that the needle has the Teflon disk as shown below. If the syringe barrel does not have the Teflon disk, use the instructions in the syringe box to wrap the needle.



- 4 Slide the spring and the cap over the needle.
- 5 Insert the needle into the syringe barrel.
- 6 Screw the cap back on the syringe barrel.

To Replace the Fused Silica Needle in a Syringe for the COC Inlet

NOTE

The fused silica needle and syringe are only used with the cooling tower and duckbill septum for manual, on-column injections onto 200- μ m columns.


1 Gather the following:

- New fused silica syringe needle (See “Consumables and Parts for the COC Inlet” on page 86.)
- Solvent

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 2 Loosen the retaining nut and remove the old needle.
- 3 Hold the syringe vertically and insert the new fused silica needle so it is visible inside the syringe barrel. If the needle cannot be inserted into the syringe barrel, the Teflon ferrule may be blocked. You may need to replace the ferrule.
- 4 Push the plunger down until it bottoms. The needle will be flush with the plunger end.



- 5 Finger-tighten the retaining nut. Pull the needle gently to be sure the Teflon ferrule has formed a tight seal with the needle. Tighten the retaining nut further, if necessary.


- 6 Loosen the retaining nut just enough so the needle is again free.
- 7 Depress the syringe plunger slowly until it pushes the needle to the end of the barrel, then finger-tighten the retaining nut.
- 8 Use a solvent to rinse the syringe and check for leaks or blocks. Leaks may be fixed by further tightening the retaining nut. Blocks or serious leaks require repeating this procedure.

To Bakeout Contaminants from the COC Inlet

- 1** Set the column flow to the normal operating value, or set the capillary column gas velocity to 30 cm/s.
- 2** Purge the column with carrier flow for at least 10 minutes before heating the oven.
- 3** Set the inlet mode to **Oven Track**.
- 4** If the column is attached to the detector, set the detector 25 °C above normal operating temperature.

If the column is not attached to the detector, cap the detector fitting.

- 5** Set the column oven 25 °C above the GC method final oven temperature to bake contaminants out of the inlet. Do not exceed the column manufacturer's maximum temperature limit.
- 6** Bakeout for 30 minutes or until the detector baseline is free of contamination peaks.



6 Maintaining the PTV Inlet

Consumables and Parts for the PTV Inlet	108
Exploded Parts View of the PTV Inlet	110
To Install a Capillary Column with the PTV Inlet	111
To Clean the Septumless Head on the PTV Inlet	114
To Replace the Septumless Head Teflon Ferrule on the PTV Inlet	117
To Change the Septum on the PTV Inlet	119
To Clean the Septum Seat in the Septum Head Assembly of the PTV Inlet	121
To Change the Liner on the PTV Inlet	123
To Replace the Inlet Adapter for the PTV Inlet	126
To Replace the Filter in the Split Vent Line	128
To Bakeout Contaminants from the PTV Inlet	129



Consumables and Parts for the PTV Inlet

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 15 PTV liners and ferrules

Description	Part number
Single baffle, 2-mm id, 180- μ L, deactivated, glass wool	5183-2038
Single baffle, 2-mm id, 200- μ L, deactivated	5183-2036
Multi-baffle, 1.5-mm id, 150- μ L, deactivated	5183-2037
Fritted glass, 1.5-mm id, 150- μ L, deactivated	5183-2041
Graphpak 3D ferrules for liner	5182-9749 (5/pk)
Installation tool for 3D ferrules	G2617-80540

Table 16 Other consumables and parts for the PTV inlet


Description	Part number
Syringe, 5- μ L, 23-gauge fixed-needle	9301-0892
Syringe, 10- μ L, 23-gauge fixed-needle	9301-0713
Syringe, 50- μ L, 23-gauge fixed-needle, for large volume injections	5183-0318
Syringe, 100- μ L, 23-gauge fixed-needle, for large volume injections	5183-2058
Split trap vent replacement kit (2 filters and 4 O-rings)	G1544-80530
Septumless head	
Septumless head	G2617-60507
Teflon ferrules (needle seal)	5182-9748 (10/pk)
Septumless head rebuild kit (includes Viton seal, Kalrez seal, and pressure spring)	5182-9747
Carrier gas tube for septumless head	G2617-80550
Ferrule, 1/16-inch Teflon, for septumless head carrier gas tube	0100-1375
Septum head	
Merlin Microseal septum (high-pressure)	5182-3444
11-mm septa, red	5181-1263 (50/pk)

Table 16 Other consumables and parts for the PTV inlet (continued)

Description	Part number
Column adapter parts	
Silver seal	5182-9763 (5/pk)
Graphpak 2M inlet adapter, 0.20-mm id*	5182-9754
Graphpak 2M inlet adapter, 0.25- to 0.33-mm id*	5182-9761
Graphpak 2M inlet adapter, 0.53-mm id*	5182-9762
Ferrules for Graphpak 2M inlet, 0.20-mm id	5182-9756 (10/pk)
Ferrules for Graphpak 2M inlet, 0.25-mm id	5182-9768 (10/pk)
Ferrules for Graphpak 2M inlet, 0.32-mm id	5182-9769 (10/pk)
Ferrules for Graphpak 2M inlet, 0.53-mm id	5182-9770 (10/pk)
Split nut for Graphpak adapter	5062-3525

* Includes (1) adapter, (1) silver seal, and (1) split column nut.

Exploded Parts View of the PTV Inlet



To Install a Capillary Column with the PTV Inlet

- 1 Gather the following (see “Consumables and Parts for the PTV Inlet” on page 108):
 - Column
 - Graphpak 2M ferrule
 - Column nut
 - Column cutter
 - Septum
 - Isopropanol
 - Lab tissue
 - Metric ruler
 - 5-mm and 6-mm wrenches
 - Typewriter correction fluid or a marker
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.


- 3 Before installing the column, install the correct column adapter. (See “To Replace the Inlet Adapter for the PTV Inlet” on page 126.)

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Place the column on the hanger with the ends pointing up and the label to the front.
- 5 Place a Graphpak 2M ferrule onto the column, with the graphite facing up towards the inlet.

- 6** Score the column using a glass scribing tool. The score must be square to ensure a clean break.



- 7** Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain that there are no burrs or jagged edges.



- 8** Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 9** Position the column so it extends 17 mm above the end of the ferrule. Mark the column behind the ferrule with typewriter correction fluid or a marker. Slide the nut over the column.



- 10** Insert the column into the adapter and finger-tighten the column nut. Looking through the slot in the nut, adjust the column until the mark is correctly positioned below the Graphpak 2M ferrule.
- 11** Tighten the column nut an additional 1/8 to 1/4 turn with a wrench. Do not overtighten.
- 12** Configure the new column.
- 13** Condition the column per the manufacturer's recommendation. (See "To Condition a Capillary Column" on page 20.)
- 14** Install the column into the detector. See:
- "To Install a Capillary Column in the FID" on page 155
 - "To Install a Capillary Column in the TCD" on page 182

- “To Install a Capillary Column in the uECD” on page 198
 - “To Install a Capillary Column Adapter in the FPD” on page 233
 - “To Install a Capillary Column in the NPD” on page 214
- 15** After the column is installed at both inlet and detector, establish a flow of carrier gas and purge as recommended by the column manufacturer.
- 16** Restore the analytical method.
- For FPD, immediately turn off the flame.
 - For NPD, immediately set the bead voltage to 0.0.
- 17** After the GC becomes ready, wait 10 minutes then ignite the detector flame or adjust offset on the NPD bead.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- 18** Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.


To Clean the Septumless Head on the PTV Inlet

- 1 Gather the following:
 - Syringe with 23-gauge needle (See “Consumables and Parts for the PTV Inlet” on page 108.)
 - Septumless head rebuild kit
 - Hexane
 - Clean, lint-free gloves
 - 5/16-inch wrench
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.
- 3 Disconnect the carrier gas line.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

- 4 Unscrew and remove the septumless head assembly from the inlet.



- 5 Unscrew the sealing element from the head assembly and carefully remove the Viton seal and the pressure spring.



- 6 Unscrew the guide cap from the head and remove the Teflon ferrule.
- 7 Carefully insert a syringe with a 23-gauge needle into the head to press the valve body and Kalrez seal slightly out of the head.
- 8 Gently tap the head on a soft, smooth surface so that the valve body falls out completely or slips out far enough to grasp with fingers.
- 9 Remove the Kalrez seal from the valve body.
- 10 Clean all components in hexane.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 11 Replace the Teflon ferrule. (See “To Replace the Septumless Head Teflon Ferrule on the PTV Inlet” on page 117.)
- 12 Wearing clean lint-free gloves, reassemble the head in reverse order. Be sure that the seals and the pressure spring are not damaged.
- 13 Finger-tighten the septumless head, then tighten an additional 1/8 turn with a wrench.
- 14 Reconnect the carrier gas line.
- 15 Check for leaks; if necessary, slightly tighten the guide cap with the syringe needle inserted.
 - If the head leaks with the syringe inserted, replace the Teflon ferrule.

6 Maintaining the PTV Inlet

- If the head leaks without the syringe inserted, replace the Kalrez and Viton seals.

16 Restore the analytical method.


To Replace the Septumless Head Teflon Ferrule on the PTV Inlet

- 1 Gather the following:
 - Syringe with 23-gauge needle (See “Consumables and Parts for the PTV Inlet” on page 108.)
 - Replacement Teflon ferrule
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

- 3 Unscrew the guide cap from the septumless head and remove the Teflon ferrule.
- 4 Push the guide cap and replacement ferrule over the syringe needle so that at least 10 mm of the needle tip is exposed.



- 5 Guide the end of the needle into the septumless head until the ferrule meets the septumless head.
- 6 Loosely install the guide cap.
- 7 If not configured, configure the column.
- 8 Put the inlet into **Splitless** mode.
- 9 Set the column flow to 5 mL/min and the purge flow to 60 mL/min.
- 10 After the inlet pressurizes, press [**Prep Run**] twice.

6 Maintaining the PTV Inlet

- 11 Observe the inlet **Total Flow**. Tighten the guide cap until the **Total Flow** stops decreasing (typically near 8 mL/min).
- 12 Remove the syringe from the inlet and press [**Stop**].
- 13 Restore the analytical method.


To Change the Septum on the PTV Inlet

- 1 Gather the following:
 - Replacement septum. (See “Consumables and Parts for the PTV Inlet” on page 108.)
 - 5/8-inch wrench
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

- 3 Remove the septum retainer nut or Merlin cap. If the septum head begins to turn during removal, support it manually while removing the cap.
- 4 Use tweezers to remove the septum or Merlin Microseal from the retainer nut. Do not gouge or scratch the interior of the septum head.
- 5 Firmly press the new septum or Merlin Microseal into the fitting. The metal parts side of the Merlin Microseal should face down (toward the oven).




- 6 Replace the septum retainer nut or Merlin cap and finger-tighten. Tighten the septum retainer nut until the C-ring is about 1 mm above the nut.

CAUTION

Overtightening the septum nut can cause contamination.

6 Maintaining the PTV Inlet



7 Restore the analytical method.

To Clean the Septum Seat in the Septum Head Assembly of the PTV Inlet

- 1 Gather the following:
 - Replacement septum (See “Consumables and Parts for the PTV Inlet” on page 108.)
 - 5/8-inch wrench
 - Tweezers
 - Compressed, filtered, dry air or nitrogen
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Remove the septum retainer nut or Merlin cap. If the septum head begins to turn during removal, support it manually while removing the cap.
- 4 Unscrew the septum head assembly from the inlet and move it up and away from the inlet.
- 5 Use tweezers to remove the septum or Merlin Microseal from the retainer nut. Do not gouge or scratch the interior of the septum head.
- 6 Scrub the residue from the retainer nut and septum holder with a small piece of rolled-up steel wool and tweezers. Do not do this over the inlet.
- 7 Use compressed air or nitrogen to blow away the pieces of steel wool and septum.
- 8 Replace the septum head assembly on the inlet. Finger-tighten the septum head and then tighten an additional 1/2 turn with a wrench.
- 9 Firmly press the new septum or Merlin Microseal into the fitting. (See “To Change the Septum on the PTV Inlet” on page 119.)

6 Maintaining the PTV Inlet

- 10** Replace the septum retainer nut or Merlin cap and finger-tighten. (See “To Change the Septum on the PTV Inlet” on page 119.)
- 11** Restore the analytical method.

To Change the Liner on the PTV Inlet

- 1 Gather the following:
 - Installation tool for 3D ferrules (See “Consumables and Parts for the PTV Inlet” on page 108.)
 - Assembly tool (part number G2617-80540)
 - Replacement liner
 - Graphpak 3D ferrule
 - 5/16-inch wrench
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Remove the head from the inlet:
 - For septumless head, disconnect the carrier gas line and unscrew and remove the septumless head assembly from the inlet.
 - For septum head, loosen the septum head assembly from the inlet. Lift the head to clear the inlet and push to either side. Do not bend the 1/16-inch lines too much.
- 4 Grasp the liner by the Graphpak 3D ferrule. Remove the liner from the inlet.



- 5** Unscrew the assembly tool into two pieces: the ferrule guide and the compression fitting.



- 6** Slide the compression fitting onto the longer, straight end of the new liner with the threads pointing toward the end of the liner.
- 7** Place a Graphpak 3D ferrule on the same end of the liner with the recessed graphite end towards the compression fitting. Slide the ferrule so that about 2 mm of the liner is exposed beyond the ferrule.
- 8** Slide the compression fitting up to meet the ferrule. Finger-tighten the ferrule guide onto the compression fitting.
- 9** Unscrew and remove the ferrule guide.
- 10** Slide the compression fitting off the other end of the liner. The ferrule should now be set with 1 mm of the liner exposed. Check that the graphite within the ferrule is flush with the top of the metal collar.

- 11 Insert the glass liner into the inlet from above until the unpacked side of the ferrule rests on the top of the inlet.
- 12 Replace the head:
 - For septumless head, screw the head onto the inlet and tighten 1/8 turn past finger-tight with a wrench. Reconnect the carrier gas line.
 - For septum head, align the head with the inlet and manually engage the free-spinning nut to the inlet. Tighten 1/2 turn past finger-tight with a wrench.
- 13 Check all connections for leaks. If necessary, tighten them again by hand.
- 14 Restore the analytical method.

To Replace the Inlet Adapter for the PTV Inlet

- 1 Select from the following list an adapter with the smallest hole diameter that will accept the column. The adapter number is stamped on the side of the adapter. (Also see “Consumables and Parts for the PTV Inlet” on page 108.)

Table 17 PTV inlet adapters

Column id	Inlet adapter number*	Part number
0.20 mm	31	5182-9754
0.25 to 0.33 mm	45	5182-9761
0.53 mm	70	5182-9762

* Includes (1) adapter, (1) silver seal, and (1) split column nut.

- 2 Gather the following:
 - Replacement adapter
 - Replacement silver seal
 - 6-mm wrench
 - 5-mm wrench
 - Lint-free gloves
- 3 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Unscrew the column nut from the adapter. Remove the nut and the column from the inlet.



- 5** Remove the inlet adapter and discard the old silver seal.
- 6** Insert a new silver seal into the adapter and finger-tighten the adapter onto the inlet. Tighten an additional 1/16 to 1/8 turn with a wrench; overtightening will damage the inlet.
- 7** Install the column. (See “To Install a Capillary Column with the PTV Inlet” on page 111.)
- 8** Check the adapter for leaks.
- 9** Restore the analytical method.

To Replace the Filter in the Split Vent Line

- 1 Gather the following:
 - New filter cartridge. (See “Consumables and Parts for the PTV Inlet” on page 108.)
 - T-20 Torx screwdriver
- 2 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

The split vent trap may contain residual amounts of any samples or other chemicals you have injected into the GC. Follow your company’s safety procedures for handling these types of substances while replacing the trap filter cartridge.

- 3 Remove the plastic pneumatics cover (top, back of GC).
- 4 Lift the filter trap assembly from the mounting bracket and unscrew the filter trap assembly.
- 5 Remove the old filter cartridge and two O-rings.



- 6 Verify the new O-rings are seated properly on the new filter cartridge.
- 7 Install the new filter cartridge then reassemble the trap.
- 8 Place the filter trap assembly in the mounting bracket.
- 9 Check for leaks.
- 10 Restore the analytical method.

To Bakeout Contaminants from the PTV Inlet

- 1 Put the inlet into split mode.
- 2 Set the column flow to the normal operating value, or set the capillary column gas velocity to 30 cm/s.
- 3 Set the inlet split vent flow to 200 mL/min.
- 4 Purge the column with carrier flow for at least 10 minutes before heating the oven.
- 5 If the column is attached to the detector, set the detector 25 °C above normal operating temperature.


WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

If the column is not attached to the detector, cap the detector fitting.

- 6 Set the inlet temperature to 300 °C or 25 °C above the normal operating temperature to bakeout contaminants from the inlet, mostly through the split vent.
- 7 Set the column oven 25 °C above the GC method final oven temperature to bake contaminants from the column. Do not exceed the column manufacturer's maximum temperature limit.
- 8 Bakeout for 30 minutes or until the detector baseline is free of contamination peaks.

6 Maintaining the PTV Inlet



7 Maintaining the VI

Consumables and Parts for the VI	132
Exploded Parts View of the VI	134
To Install a Capillary Column with the VI	135
To Attach a Sample Transfer Line to the VI	138
To Remove the VI Interface	139
To Clean the VI	140
To Install the VI Interface	141
To Replace the Filter in the Split Vent Line	142
To Bakeout Contaminants from the VI Inlet	143



Consumables and Parts for the VI

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 18 Parts for the VI

Description	Part number
Clamping plate	G2319-20540
Volatiles interface	G2319-60505
Long column nut (65 mm)	G3504-20504
Nut, for transfer, pressure sensing, or split vent line	19258-20830
Ferrule, for transfer, pressure sensing, or split vent line	19258-20870
Split vent trap replacement kit (2 filters and 4 O-rings)	G1544-80530


Table 19 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292

Table 19 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Exploded Parts View of the VI



To Install a Capillary Column with the VI

- 1 Gather the following:
 - Long (65 mm) column nut
 - Column
 - Ferrule
 - Column cutter
 - Isopropanol
 - Septum
 - 1/4-inch wrench
 - Lab tissue
 - Metric ruler
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.


- 3 Place the column on the hanger with the ends pointing up and the label to the front.
- 4 Place a septum, long capillary column nut, and ferrule on the column.

Use a long column nut. (See “Consumables and Parts for the VI” on page 132.)

If you are using a standard column nut, you must remove the interface. For this reason it is recommended that you use the long column nut. (See “To Remove the VI Interface” on page 139.)



- 5** Score the column using a glass scribing tool. The score must be square to ensure a clean break.



- 6** Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain that there are no burrs or jagged edges.



- 7** Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 8** Position the column so it extends 6 mm above the end of the ferrule. Slide the septum up the column to hold the column nut at this fixed position.



- 9 Insert the column into the interface and finger-tighten the column nut.
- 10 Adjust the column (*not* the septum) position until the septum is snug against the bottom of the nut.
- 11 Tighten the column nut an additional 1/4 to 1/2 turn with a wrench so that the column cannot be pulled from the fitting with gentle pressure.
- 12 Configure the new column.
- 13 With the sample transfer line attached and the column attached to the inlet and detector, establish a flow of carrier gas through the transfer line. Purge as recommended by the column manufacturer.
- 14 Condition the column per the manufacturer's recommendation. (See "To Condition a Capillary Column" on page 20.)
- 15 Install the column into the detector. See:
 - "To Install a Capillary Column in the FID" on page 155
 - "To Install a Capillary Column in the TCD" on page 182
 - "To Install a Capillary Column in the uECD" on page 198
 - "To Install a Capillary Column Adapter in the FPD" on page 233
 - "To Install a Capillary Column in the NPD" on page 214
- 16 After the column is installed at both inlet and detector, establish a flow of carrier gas and purge as recommended by the column manufacturer.
- 17 Restore the analytical method.
 - For FPD, immediately turn off the flame.
 - For NPD, immediately set the bead voltage to 0.0.
- 18 After the GC becomes ready, wait 10 minutes then ignite the detector flame or adjust offset on the NPD bead.


WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.


- 19 Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.

To Attach a Sample Transfer Line to the VI

- 1 Gather one 7/16-inch, two 5/16-inch, and one 7-mm wrenches.
- 2 Attach the transfer line from the gas sampler by finger-tightening the preattached transfer line nut and ferrule. Tighten an additional 1/4 turn with a wrench.



If the transfer line is from a G1900A Purge and Trap, install the transfer line support nut assembly up and inside the metal sleeve of the heated line assembly to prevent damage to the fused silica line.



- 3 With the column installed, establish a flow of carrier gas through the transfer line and check for leaks. If the transfer line nut leaks, tighten an additional 1/8 turn with a wrench. Purge as recommended by the column manufacturer.
- 4 Heat the interface to operating temperature.
- 5 Retighten the fittings, if necessary.


To Remove the VI Interface

- 1 Gather the following:
 - 1/4-inch and 7-mm wrench
 - T-20 Torx screwdriver
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The interface may be hot enough to cause burns. If the interface is hot, wear gloves to protect your hands.

- 3 Remove the column.
- 4 Remove the transfer line by loosening the nut, then lift it from the interface.
- 5 Loosen the five inlet cover screws and remove the inlet cover.
- 6 Remove the clamping plate from the interface by loosening the captive screw.



- 7 Lift the interface out of the heater block.



To Clean the VI

- 1 Gather the following:
 - 1/4-inch and 7-mm wrench
 - T-20 Torx screwdriver
 - Lint-free gloves
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.
- 3 Remove the interface. (See “To Remove the VI Interface” on page 139.)


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

The split vent trap and line may contain residual amounts of any samples or other chemicals you have injected into the GC. Follow your company’s safety procedures for handling these types of substances.

- 4 Remove the split vent and pressure sensing lines by loosening the nuts.




CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.


- 5 Clean the interface using an ultrasonic bath. Sonicate twice, then rinse and air dry.
- 6 Inspect the split vent line. If clogged, contact Agilent for service.
- 7 Install the interface. (See “To Install the VI Interface” on page 141.)

To Install the VI Interface

- 1 Gather the following:
 - 1/4-inch and 7-mm wrench
 - T-20 Torx screwdriver
- 2 Attach the split vent line and pressure sensing lines and finger-tighten the nuts. Tighten an additional 1/4 turn with a wrench.



- 3 Place the interface into the heater block and reseal the tubing as needed.
- 4 Install the clamping plate and tighten the screw.



- 5 Install the inlet cover. Be sure the cover does not damage any tubing.
- 6 Attach the sample transfer line. (See "To Attach a Sample Transfer Line to the VI" on page 138.)
- 7 Install the column. (See "To Install a Capillary Column with the VI" on page 135.)

To Replace the Filter in the Split Vent Line

- 1 Gather the following:
 - New filter cartridge. (See “Consumables and Parts for the VI” on page 132.)
 - T-20 Torx screwdriver
- 2 Prepare the inlet for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.

WARNING

The split vent trap may contain residual amounts of any samples or other chemicals you have injected into the GC. Follow your company’s safety procedures for handling these types of substances while replacing the trap filter cartridge.

- 3 Remove the plastic pneumatics cover (top, back of GC).
- 4 Lift the filter trap assembly from the mounting bracket and unscrew the filter trap assembly.
- 5 Remove the old filter cartridge and two O-rings.



- 6 Verify the new O-rings are seated properly on the new filter cartridge.
- 7 Install the new filter cartridge then reassemble the trap.
- 8 Place the filter trap assembly in the mounting bracket.
- 9 Check for leaks.
- 10 Restore the analytical method.

To Bakeout Contaminants from the VI Inlet


- 1 Put the inlet into split mode.
- 2 Set the column flow to the normal operating value, or set the capillary column gas velocity to 30 cm/s.
- 3 Set the inlet split vent flow to 200 mL/min.
- 4 Purge the column with carrier flow for at least 10 minutes before heating the oven.
- 5 If the column is attached to the detector, set the detector 25 °C above normal operating temperature.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

If the column is not attached to the detector, cap the detector fitting.

- 6 Set the inlet temperature to 300 °C or 25 °C above the normal operating temperature to bakeout contaminants from the inlet, mostly through the split vent.
- 7 Set the column oven 25 °C above the GC method final oven temperature to bake contaminants from the column. Do not exceed the column manufacturer's maximum temperature limit.
- 8 Bakeout for 30 minutes or until the detector baseline is free of contamination peaks.



8 Maintaining the FID

Consumables and Parts for the FID	146
Exploded Parts Views of the FID	149
Selecting an FID jet	151
To Attach a Capillary Column Adapter on an Adaptable FID	153
To Install a Capillary Column in the FID	155
To Replace the FID Collector Assembly	158
To Replace an FID Jet	160
To Perform Maintenance on the FID Collector Assembly	163
To Check the FID Leakage Current	171
To Check the FID Baseline	172
To Install the FID Insulation Cup Assembly (Adaptable FID Only)	173
To Install the Optional FID PTFE Chimney Insert	175
To Bakeout the FID	176



Consumables and Parts for the FID

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 20 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Table 21 FID parts and subassemblies

Description	Part number/quantity
Screw, M4 × 25 mm, Torx, T20	0515-2712 (3/pk)
PTFE chimney (optional)	19231-21050
Collector assembly	G1531-60690
FID/NPD capillary column adapter	19244-80610
FID/NPD 1/8-inch packed column adapter	19231-80520
FID/NPD 1/4-inch packed column adapter	19231-80530
Insulation	19234-60715 (3/pk)
Insulation cup assembly	19234-60700
Nut, 1/4-inch, brass, for packed column adapters	5180-4105 (10/pk)
Ferrule, Vespel, 1/4-inch, for packed column adapters	5080-8774 (10/pk)

Table 22 Jets for capillary adaptable fittings

Jet type	Part number	Jet tip id	Length
Capillary	19244-80560	0.29 mm (0.011 inch)	61.5 mm
Capillary, high-temperature (use with simulated distillation)	19244-80620	0.47 mm (0.018 inch)	61.5 mm
Packed	18710-20119	0.46 mm (0.018 inch)	63.6 mm
Packed, wide-bore (use with high-bleed applications)	18789-80070	0.76 mm (0.030 inch)	63.6 mm

Table 23 Jets for capillary optimized fittings

Jet type	Part number	Jet tip ID	Length
Capillary	G1531-80560	0.29 mm (0.011 inch)	48 mm
High-temperature (use with simulated distillation)	G1531-80620	0.47 mm (0.018 inch)	48-mm


Table 24 FID collector assembly parts

Description	Part number/quantity
Screw, M4 × 25 mm, Torx, T20	0515-2712 (3/pk)
Collector assembly	G1531-60690
Collector nut	19231-20940





Table 24 FID collector assembly parts (continued)

Description	Part number/quantity
Spring washer	3050-1246
Ignitor castle	19231-20910
Ignitor castle, Hastelloy	19231-21060
Upper/lower collector insulator	G1531-20700
Collector body	G1531-20690
Collector body, Hastelloy	G1531-21090
Spanner nut (collector)	19231-20980
Collector mount	G1531-20550
Collector housing	G1531-20740
Gasket	5180-4165 (12/pk)
Ignitor (glow plug) assembly with O-ring	19231-60680

Exploded Parts Views of the FID




8 Maintaining the FID



Selecting an FID jet

Open the oven door and locate the column connection fitting at the base of the detector. It will look like either a capillary optimized fitting or an adaptable fitting.

Capillary optimized fitting



Adaptable fitting



- If you have an application that tends to clog the jet, select a jet with a wider tip id.
- When using packed columns in high column-bleed applications, the jet tends to clog with silicon dioxide.
- In simulated distillation applications, the high-boiling hydrocarbons tend to clog the jet.

For capillary optimized fittings, select a jet from Table 25.

Table 25 Jets for capillary optimized fittings

Figure 1 ID	Jet type	Part number	Jet tip id	Length
1	Capillary	G1531-80560	0.29 mm (0.011 inch)	48 mm
2	High-temperature (use with simulated distillation)	G1531-80620	0.47 mm (0.018 inch)	48 mm



Figure 1 Capillary optimized jets

For adaptable fittings, select a jet from Table 26.

Table 26 Jets for capillary adaptable fittings

Figure 2 ID	Jet type	Part number	Jet tip id	Length
1	Capillary	19244-80560	0.29 mm (0.011 inch)	61.5 mm
2	Capillary, high-temperature (use with simulated distillation)	19244-80620	0.47 mm (0.018 inch)	61.5 mm
3	Packed	18710-20119	0.46 mm (0.018 inch)	63.6 mm
4	Packed, wide-bore (use with high-bleed applications)	18789-80070	0.76 mm (0.030 inch)	63.6 mm



Figure 2 Capillary adaptable jets

To Attach a Capillary Column Adapter on an Adaptable FID

- 1 Gather the following materials:
 - Adapter (See “Consumables and Parts for the FID” on page 146.)
 - 1/4-inch brass nut
 - 1/4-inch Vespel/graphite ferrule
 - Column cutter
 - 1/4-inch wrench
 - 9/16-inch open-end wrench
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

CAUTION


Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble the nut and ferrule onto the adapter.



- 4 Insert the adapter straight into the detector base as far as possible.
- 5 Hold the adapter in this position and finger-tighten the nut.

8 Maintaining the FID



6 Tighten an additional 1/4 turn with a wrench.

To Install a Capillary Column in the FID

- 1 Gather the following materials (see “Consumables and Parts for the FID” on page 146.):
 - Column
 - Ferrule(s)
 - Column nut
 - Column cutter
 - 1/4-inch open-end wrench
 - Septum
 - Isopropanol
 - Lab tissue
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 If using the adaptable detector, verify that the adapter is installed. (See “To Attach a Capillary Column Adapter on an Adaptable FID” on page 153.)

Capillary optimized fitting




Adaptable fitting




CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Place a septum, capillary column nut, and ferrule on the column.



- 5 Score the column using a glass scribing tool. The score must be square to ensure a clean break.



- 6 Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.




- 7 Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.

- 8 Install the capillary column.


If the column id is greater than 0.1 mm:

- a Gently insert the column into the detector until it bottoms; do not attempt to force it further.
- b Finger-tighten the column nut, then withdraw the column about 1 mm. Tighten the nut an additional 1/4 turn with a wrench.

If the column id is 0.1 mm or less position the column so it extends above the ferrule by 48 mm (*capillary optimized fitting*) or 68 mm (*adaptable fitting*). Slide the septum up to hold the column nut and ferrule at this fixed position.



- c Insert the column into the detector. Slide the nut and ferrule up the column to the detector base. Finger-tighten the column nut until it grips the column.
- d Adjust the column (*not* the septum) position so that the septum is even with the bottom of the column nut. Tighten the nut an additional 1/4 turn with a wrench.



To Replace the FID Collector Assembly

1 Gather the following:


- New FID collector assembly. (See “Consumables and Parts for the FID” on page 146.)
- T-20 Torx screwdriver
- 1/4-inch nut driver
- Tweezers
- Lint-free gloves

CAUTION


To avoid contaminating the FID, wear clean, lint-free gloves when handling the collector assembly.

2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

3 Disconnect the ignitor cable assembly.



4 Remove the three screws holding the collector assembly to the mounting pallet.




CAUTION

This step exposes the interconnect spring. Be careful not to touch or disfigure the spring while working on the FID. Any dirt or bending will reduce the sensitivity of your detector.

- 5 Lift and remove the assembly from the pallet.



- 6 Remove the ignitor cable assembly from the new collector assembly, if present.
- 7 Remove any protective caps from the new collector assembly, if present.
- 8 Lower the new collector assembly into the housing.
- 9 Insert the three screws and tighten (to 18 inch-pounds).



- 10 Connect the ignitor extension cable.
- 11 Verify assembly:
 - a Check the FID leakage current. (See “To Check the FID Leakage Current” on page 171.)
 - b Check the FID baseline. (See “To Check the FID Baseline” on page 172.)

To Replace an FID Jet

- 1 Gather the following:
 - Replacement jet (See “Selecting an FID jet” on page 151.)
 - T-20 Torx screwdriver
 - 1/4-inch nut driver
 - Tweezers
 - Compressed, filtered, dry air or nitrogen
 - Solvent that will clean the type of deposits in your detector
 - Clean cloth
 - Cotton swab
 - Lint-free gloves
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 If installed, remove the capillary column from the detector.

CAUTION


Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Remove the FID collector assembly and place it on a clean cloth. (See “To Replace the FID Collector Assembly” on page 158.)
- 5 Locate the jet inside the housing.

**CAUTION**

Handle the clean or new jet only with tweezers, or wear gloves.

- 6 Loosen the jet, then lift it out of the housing with tweezers.





- 7 Clean the detector base cavity using solvent, a swab, and compressed air or nitrogen.
- 8 Use tweezers to lower the new jet into the housing.

CAUTION

Do not overtighten the jet! Overtightening may permanently deform and damage the jet, the detector base, or both. The torque specification is 10 inch-pounds.

- 9 Carefully screw the jet into the housing. Tighten 1/6-turn past finger-tight (1/6-turn is one “flat” on a typical screwdriver handle, or the jet head).



- 10 Install the collector assembly. (See “To Replace the FID Collector Assembly” on page 158.)
- 11 Attach the capillary column to the detector.
 - a Install the column in the detector. (See “To Install a Capillary Column in the FID” on page 155.)
 - b After the column is installed at both inlet and detector, establish a flow of carrier gas and purge as recommended by the column manufacturer.
 - c Check the FID leakage current. (See “To Check the FID Leakage Current” on page 171.)
 - d Bakeout the detector. (See “To Bakeout the FID” on page 176.)
 - e Restore the analytical method.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

- f Allow the oven, inlet, and detector to equilibrate at operating temperature, then retighten the fittings.
- 12 Check the FID baseline. (See “To Check the FID Baseline” on page 172.)

To Perform Maintenance on the FID Collector Assembly

NOTE

Perform only the steps and gather only the parts that apply to the desired maintenance task(s).

1 Gather the following:

- Replacement ignitor assembly (See “Consumables and Parts for the FID” on page 146.)
- Replacement ignitor castle
- Two collector insulators
- Collector
- Spring washer
- Gasket
- T-20 Torx screwdriver
- 1/4-inch nut driver
- Tweezers
- 5/16-inch wrench
- Lint-free gloves
- Clean cloth

CAUTION


To avoid contaminating the FID, wear clean, lint-free gloves when handling the collector assembly.

2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.


- 3** Remove the FID ignitor.
 - a** Disconnect the ignitor cable assembly.



- b** Loosen the ignitor with a wrench.



- c** Turn the nut counterclockwise by hand. Remove the ignitor and copper washer.
- 4** If replacing only the FID ignitor assembly with O-ring, skip to step 16 for assembly.
- 5** Remove the three screws that hold the collector mount to the FID thermal strap.



CAUTION

This step exposes the interconnect spring. Be careful not to touch or disfigure the spring while working on the FID. Any dirt or bending will reduce the sensitivity of your detector.

- 6 Remove the collector assembly. Place it on a clean cloth for additional disassembly.



- 7 Remove the gasket from the bottom of the assembly, if necessary.
- 8 Remove the FID ignitor castle.
- a Loosen the collector nut.
 - b Remove the collector nut and the spring washer.



- c Lift the castle out of the collector housing. When removing the castle, some of the collector parts may be attached. Set these on a clean cloth to protect from scratches or dirt.



9 If only replacing the FID castle, skip to step 15 for reassembly.

10 Remove the collector and insulators.

- a** If needed, remove the collector and upper insulator from the FID housing. The lower insulator may come out with the collector, but often remains in the FID housing. Place the parts on a clean cloth.




- b** Remove the lower insulator with tweezers and place the parts on a clean cloth.



- 11 Remove the collector housing from the mount, if necessary.
- 12 Use tweezers to remove the gasket from the bottom of the housing.

The collector assembly is now completely disassembled.
Reassemble as follows:

- 13 Use tweezers to install a new gasket onto the housing, being sure that it lays flat on the brass surface.



- 14 Install the collector insulators.
 - a Insert one of the insulators into the base of the housing. Seat the insulator with the flat surface facing out of the housing.
 - b Insert the long end of collector into the housing and lower insulator.



- c** Insert the other insulator onto the top of the collector, with the flat surface facing towards the housing.



15 Install the FID ignitor castle.


- a** Orient the castle so that the threaded hole for the ignitor faces toward the electronics.



- b** Insert the FID castle into the collector housing.
- c** Install the spring washer over the castle.



- d** Install the collector nut over the castle and tighten firmly. The seal should be airtight. Maintain the orientation of the ignitor hole with the base as shown below.




16 Install the FID ignitor.

- a** Insert the ignitor and copper seal into the threaded hole of the castle. Keep the mating threads clean.




- b** Tighten the ignitor with a wrench. Ignition requires a good electrical contact that is free of any dirt.



17 Lower the collector assembly into the housing.

18 Insert the three screws and tighten (to 18 inch-pounds).



19 Connect the ignitor extension cable.

20 Verify assembly:

- a** Check the FID leakage current. (See “To Check the FID Leakage Current” on page 171.)
- b** Bakeout the detector. (See “To Bakeout the FID” on page 176.)
- c** Check the FID baseline. (See “To Check the FID Baseline” on page 172.)

To Check the FID Leakage Current

- 1 Load the analytical method.
 - Make sure flows are acceptable for ignition.
 - Heat the detector to operating temperature or 300 °C.
- 2 Turn off the FID flame.
- 3 Verify that the FID electrometer is on.
- 4 Press **[Front Detector]** or **[Back Detector]**, then scroll to **Output**.
- 5 Verify that the output is stable and < 1.0 pA.

If the output is unstable or > 1.0 pA, turn off the GC and check for proper assembly of the upper FID parts and contamination. If this contamination is confined to the detector, bakeout the FID. (See “To Bakeout the FID” on page 176.)

- 6 Turn on the flame.

To Check the FID Baseline

- 1** With the column installed, load your checkout method.
- 2** Set the oven temperature to 35 °C.
- 3** Press [**Front Detector**] or [**Back Detector**], then scroll to **Output**.
- 4** When the flame is lit and the GC is ready, verify that the output is stable and < 20 pA.

If the output is not stable or > 20 pA, the system or gas may be contaminated. If this contamination is isolated to the detector, then bakeout the FID. (See “To Bakeout the FID” on page 176.)

To Install the FID Insulation Cup Assembly (Adaptable FID Only)

- 1 Gather the following:
 - Insulation (See “Consumables and Parts for the FID” on page 146.)
 - Insulation cup assembly
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING


Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 3 Assemble the insulation in the cup. Line up the slots in the insulation with the slot in the cup.


Capillary columns should be attached to the detector before installing the cup. When attaching a packed column to the detector, cap the detector fitting before installing the cup to prevent insulation contamination of the detector.



- 4 Push the wire spring lever to the right to uncover the hole.



- 5** From inside the oven with the column installed, pass the column through the slot in the cup. Move the cup up over the detector fitting so that the cup touches the top of the oven. You should be able to see the groove in the fitting.
- 6** Release the spring into the groove of the fitting.




To Install the Optional FID PTFE Chimney Insert

WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 1 Light the FID flame.
- 2 Insert the PTFE chimney into the FID castle.

**NOTE**

When installed, the PTFE chimney insert prevents ignition.

To Bakeout the FID

WARNING

If using hydrogen as a carrier gas, turn off the hydrogen supply and cap the end of the column to prevent an oven explosion.


- 1 Bakeout the FID with the column installed or uninstalled. If uninstalled, gather the following (see “Consumables and Parts for the FID” on page 146):
 - Capillary adapter (adaptable FID only)
 - Column nut
 - No-hole ferrule
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 3 If the column is uninstalled, plug the detector connection with the capillary adapter, column nut, and no-hole ferrule.


Maintain inert carrier gas flow through the column, or remove the column from the GC.
- 4 Set the detector temperature at 350 to 375 °C.
- 5 Set normal operating flows.
- 6 Light the FID flame.
- 7 Set the oven temperature to 250 °C or 25 °C above the normal maximum operating temperature. Do not exceed the column’s temperature limit.
- 8 Hold at temperature for 30 minutes or until the baseline settles at a lower value. The baseline will typically rise, then fall to a final value lower than the initial baseline.



- 9** Restore the analytical method and allow the FID to equilibrate.
- 10** Check the FID output value. It should be lower than the first reading. If it is not, contact your Agilent service representative.

Without a column installed, a clean system baseline should be ≤ 20 pA.

- 11** If the column is not installed in the FID, install it. (See “To Install a Capillary Column in the FID” on page 155.)



9 Maintaining the TCD

Consumables and Parts for the TCD	180
To Install a Capillary Column in the TCD	182
To Install the Optional TCD Capillary Column Adapter	184
To Install a Capillary Column with the Optional TCD Capillary Column Adapter	185
To Bakeout Contaminants from the TCD	187



Consumables and Parts for the TCD

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Standard TCD column hardware

Table 27 Standard parts for attaching columns to the TCD

Column	Description	Unit	Part number
Capillary	Nut, 1/8-inch id, brass Swagelok	10/pk	5180-4103
	Back ferrule, for 0.1-mm to 0.53-mm capillary columns	10/pk	5182-3477
	Front ferrule, 0.53-mm capillary columns	10/pk	5182-9673
	Front ferrule, 0.32-mm capillary columns	10/pk	5182-9676
	Front ferrule, 0.1-mm, 0.2-mm, and 0.25-mm capillary columns	10/pk	5182-9677
	1/8-inch Swagelok plug		5180-4124
1/4-inch packed	1/4-inch packed column adapter		G1532-20710
	1/8-inch id Vespel/graphite ferrule	10/pk	0100-1332
	Nut, 1/8-inch id, brass	10/pk	5180-4103
	Ferrule, Vespel, 1/4-inch	10/pk	5080-8774
	1/4-inch id tubing nut, brass	10/pk	5180-4105
	1/8-inch Swagelok plug		5180-4124
1/8-inch packed	Ferrule, 1/8-inch Vespel/graphite	10/pk	0100-1332
	Nut, 1/8-inch id, brass	10/pk	5180-4103
	1/8-inch Swagelok plug		5180-4124

Optional TCD capillary column hardware

Table 28 Optional TCD capillary column adapter hardware

Description	Unit	Part number
Capillary adapter		G1532-80540
Ferrule, Vespel, 1/8-inch	10/pk	0100-1332
Nut, brass, 1/8-inch	10/pk	5180-4103

Table 29 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

To Install a Capillary Column in the TCD

- 1 Gather the following:
 - Front ferrule (See “Consumables and Parts for the TCD” on page 180.)
 - Back ferrule
 - Column nut
 - Column cutter
 - 7/16-inch wrench
 - Lab tissue
 - Lint-free gloves
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or inlet may be hot enough to cause burns. If either is hot, wear heat-resistant gloves to protect your hands.


WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble the ferrules and 1/8-inch brass Swagelok nut on the column.



- 4 Score the column using a glass scribing tool. The score must be square to ensure a clean break.



- 5 Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain that there are no burrs or jagged edges.



- 6 Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 7 Insert the column into the detector until it bottoms.
- 8 Slide the column nut and ferrules up the column to the detector and finger-tighten the nut.
- 9 Pull out 1 mm of column. Tighten the nut an additional 1/4 turn with a wrench or until the column does not move.


To Install the Optional TCD Capillary Column Adapter

- 1 Gather the following:
 - Capillary column adapter (See “Consumables and Parts for the TCD” on page 180.)
 - 1/4-inch and 7/16-inch wrenches
 - 1/8-inch brass nut
 - 1/8-inch Vespel ferrule
 - Lint-free gloves
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble the brass nut and ferrule onto the capillary column adapter.



- 4 Install the adapter assembly into the detector fitting and tighten finger-tight. Use a wrench to tighten until snug.



To Install a Capillary Column with the Optional TCD Capillary Column Adapter

- 1 Gather the following:
 - Ferrule (See “Consumables and Parts for the TCD” on page 180.)
 - Column cutter
 - Column nut
 - 1/4-in. and 7/16-in. wrenches
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.


WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Place a capillary column nut and ferrule on the column.




- 4 Score the column using a glass scribing tool. The score must be square to ensure a clean break.




- 5** Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.



- 6** Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 7** Gently insert the column into the detector until it bottoms. Do not attempt to force it further.



- 8** Slide the column nut and ferrule up the column to the adapter and tighten the nut finger tight.



- 9** Pull the column out 1 mm. Use a wrench to tighten the nut an additional 1/4-turn. The column should not move.

To Bakeout Contaminants from the TCD

The bakeout can be performed with the column installed or the detector capped.

CAUTION

If the column is not installed, you must turn off the TCD filament and cap the detector column fitting to prevent irreparable damage to the filament caused by oxygen entering the detector.

- 1 If the column is not installed, cap the detector.

WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 2 Turn off the TCD filament.
- 3 If the column is attached to the inlet, maintain inert carrier gas flow through the column.

WARNING


If using hydrogen as a carrier gas, turn off the hydrogen supply and cap the end of the column to prevent an oven explosion.

- 4 Set the reference gas flow rate between 20 and 30 mL/min.
- 5 Set the detector temperature to 375 °C.
- 6 Hold at 375 °C for several hours.
- 7 If the column is uninstalled, install it. (See “To Install a Capillary Column in the TCD” on page 182.)
- 8 Load the analytical method.

WARNING

Be careful! The oven or detector fittings may be hot enough to cause burns.

- 9 Allow the oven, inlet, and detector to equilibrate at operating temperature, then re-tighten the fittings.



10 Maintaining the uECD

Important Safety Information About the uECD	190
Consumables and Parts for the uECD	192
Exploded Parts View of the uECD	194
To Replace the uECD Fused Silica Indented Mixing Liner and Install the Makeup Gas Adapter	195
To Install a Capillary Column in the uECD	198
To Install the Detector Insulating Cup	200
To Bakeout the uECD	202

This section describes the routine maintenance tasks for the micro-Electron Capture Detector (uECD). For important regulatory and safety information for this detector, refer to the general information booklet (part number 5961-5664) and CD (part number G1533-91001) provided with the detector.



Important Safety Information About the uECD

The uECD contains a cell plated with ^{63}Ni , a radioactive isotope. The beta particles released at the energy level in the detector have little penetrating power—the surface layer of the skin or a few sheets of paper will stop most of them—but they may be hazardous if the isotope is ingested or inhaled. For this reason, handle the cell with care. Cap the detector inlet and outlet fittings when the detector is not in use. Never introduce corrosive chemicals into the detector. Vent detector exhaust outside the laboratory environment.

Refer to the safety documentation provided with the detector for important details about safety, maintenance, and compliance with local government regulation.

WARNING

Materials that may react with the ^{63}Ni source, either to form volatile products or to cause physical degradation of the plated film, must be avoided. These materials include oxidizing compounds, acids, wet halogens, wet nitric acid, ammonium hydroxide, hydrogen sulfide, PCBs, and carbon monoxide. This list is not exhaustive but indicates the kinds of compounds that may cause damage to ^{63}Ni detectors.

WARNING

In the extremely unlikely event that both the oven and the detector-heated zone should go into thermal runaway (maximum, uncontrolled heating in excess of 400 °C) at the same time and the detector remains exposed to this condition for more than 12 hours, take the following steps:

- 1 After turning off the main power and allowing the instrument to cool, cap the detector inlet and exhaust vent openings. Wear disposable plastic gloves and observe normal laboratory safety precautions.
- 2 Return the cell for disposal, following directions included with the License Verification Form (part number 19233-90750).
- 3 Include a letter stating the condition of abuse.

It is unlikely, even in this very unusual situation, that radioactive material will escape the cell. However, permanent damage to the ^{63}Ni plating within the cell is possible; therefore, the cell must be returned for exchange.

WARNING

Do not use solvents to clean the uECD.

WARNING

You may not open the uECD cell unless authorized to do so by your local nuclear regulatory agency. Do not disturb the four socket-head bolts. These hold the cell halves together. Removing or disturbing them is a violation of the terms of the General License and could create a safety hazard.

When handling uECDs:

- Never eat, drink, or smoke.
- Always wear safety glasses when working with or near open uECDs.
- Wear protective clothing such as laboratory jackets, safety glasses, and gloves, and follow good laboratory practices. Wash hands thoroughly with a mild nonabrasive cleaner after handling uECDs.
- Cap the inlet and outlet fittings when the uECD is not in use.
- Connect the uECD exhaust vent to a fume hood or vent it to the outside. See the latest revision of 10 CFR Part 20 (including Appendix B), or the applicable state regulation. For other countries, consult with the appropriate agency for equivalent requirements.

Agilent Technologies recommends a vent line internal diameter of 6 mm (1/4-inch) or greater. With a line of this diameter, the length is not critical.

Consumables and Parts for the uECD

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 30 uECD consumables and parts

Description	Part number/quantity
Fused silica indented mixing liner	G2397-20540
Makeup gas adapter	G2397-80520
ECD wipe test kit	18713-60050
Insulation	19234-60715 (3/pk)
Insulation cup assembly	19234-60700
Nut, 1/4-inch Swagelok adapter	5180-4105 (10/pk)
Ferrule, graphitized Vespel, 1/4-inch	5080-8774 (10/pk)
Capillary column blanking nut	5020-8294


Table 31 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292

Table 31 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Exploded Parts View of the uECD



To Replace the uECD Fused Silica Indented Mixing Liner and Install the Makeup Gas Adapter

- 1 Gather the following:
 - Fused silica indented mixing liner (See “Consumables and Parts for the uECD” on page 192.)
 - 1/4-inch Swagelok nut
 - 1/4-inch Vespel/graphite ferrule
 - 9/16-inch wrench
 - Methanol
 - Lint-free gloves
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.


- 3 Remove the insulating cup from the detector base.
- 4 Remove the column from the makeup gas adapter.

CAUTION

To prevent damage avoid flexing/bending the tubing on the makeup gas adapter.

- 5 Remove the makeup gas adapter.
 - a Loosen the adapter nut with a wrench and slide out the makeup gas adapter from the uECD. Remove the ferrule.

The makeup gas adapter will remain attached to the supply tubing and hang suspended in the oven.
 - b Adjust the adapter's position so that maintenance can be performed on the adapter easily and without obstruction.




- 6 Unscrew and remove the adapter cap.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.


- 7 Remove the fused silica indented mixing liner and inspect. Replace it if it is broken or contaminated with sample or graphite.
- 8 Ultrasonically clean the adapter cap in methanol. Clean the outer surfaces of the makeup gas adapter with methanol.
- 9 Install the fused silica indented mixing liner into the makeup gas adapter, then install the cap. The indentation on the fused silica indented mixing liner must be at the cap end of the adapter.



- 10 Place a new 1/4-inch Swagelok nut and ferrule onto the makeup gas adapter.




- 11** Slowly install the adapter straight into the detector fitting. Jiggle the adapter, if necessary, to make sure it is seated all the way into the detector fitting. Be careful not to break the column end.



- 12** Tighten the nut finger-tight and then use a 9/16-inch wrench to tighten until snug.

If the adapter is properly installed, the distance between the 1/4-inch nut and the bottom of the adapter will be 19 ± 1 mm. If the distance is 22 to 23 mm, install the adapter into the detector fitting.



- 13** Attach the column. (See “To Install a Capillary Column in the uECD” on page 198.)

To Install a Capillary Column in the uECD

- 1 Gather the following:
 - Ferrule (See “Consumables and Parts for the uECD” on page 192.)
 - Column nut
 - Septum
 - Column
 - 1/4-inch, 5/16-inch, and 9/16-inch wrenches
 - Column cutter
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 Prepare the inlets for maintenance. See “Preparing the GC for Maintenance” on page 15.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Place a septum, capillary column nut, and ferrule on the column.



- 5 Score the column using a glass scribing tool. The score must be square to ensure a clean break.




- 6 Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.



- 7 Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 8 Install the column.

If the column id is 200 μ m or more, push the column into the adapter until it stops at the indentation. Pull it back 1 to 2 mm and tighten the column nut with one 5/16-inch wrench on the adapter and another 1/4-inch wrench on the column nut.

If the id is less the 200 μ m, mark the column with a septum 70 ± 1 mm from the end. Insert column and nut into the adapter with the septum at the rear of the column nut, and tighten the column nut with one 5/16-inch wrench on the adapter and another 1/4-inch wrench on the column nut.



- 9 After heating the detector, retighten the 9/16-inch makeup adapter nut and 1/4-inch column nut.

To Install the Detector Insulating Cup

- 1 Gather the following:
 - Nut warmer insulation (See “Consumables and Parts for the uECD” on page 192.)
 - Insulation cup assembly
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.


WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 3 Install the insulation in the cup. Line up the slots in the insulation with the slot in the cup.




- 4 Push the wire spring lever to the right to uncover the hole.



- 5 With a capillary column installed to prevent insulation contamination of the detector, slide the column into the slot on the insulation cup and place the cup over the makeup gas adapter.

- 6 Slide the cup up so that the cup touches the top of the oven and you can see the groove in the makeup gas adapter.
- 7 Release the spring into the groove of the makeup gas adapter.



To Bakeout the uECD

WARNING

Detector disassembly and/or cleaning procedures other than thermal should be performed only by personnel trained and licensed appropriately to handle radioactive materials. Trace amounts of radioactive ^{63}Ni may be removed during other procedures, causing possible hazardous exposure to b- and x-radiation.

CAUTION

To prevent possible hazardous contamination of the area with radioactive material, the detector exhaust vent always must be connected to a fume hood or otherwise vented in compliance with the latest revision of 10 CFR Part 20, or with state regulations with which the Nuclear Regulatory Commission has entered into an agreement (USA only). For other countries, consult with the appropriate agency for equivalent requirements.

- 1 Gather the following:
 - Column nut and no-hole ferrule (See “Consumables and Parts for the uECD” on page 192.)
 - Blanking nut with any column ferrule
- 2 With the detector and oven at normal operating temperatures, press [**Front Det**] or [**Back Det**]. Note the value of **Output** for later comparison.
- 3 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 4 If the column maximum temperature is $< 250\text{ }^{\circ}\text{C}$, remove the column from the detector.


- 5 If the column is uninstalled, plug the detector connection with the column nut and no-hole ferrule.

Maintain inert carrier gas flow through the column, or remove the column from the GC.

- 6 Set the uECD temperature to 350 to 375 °C, the makeup gas flow to 60 mL/min, and the oven temperature to 250 °C. If the column is uninstalled, leave the oven off to protect the column.
- 7 If the column is installed in the uECD, set the oven temperature to 250 °C. If the column is uninstalled, leave the oven off to protect the column.
- 8 Allow thermal cleaning to continue for several hours and then cool the system to normal operating temperatures. The figure below shows detector output during a typical cleaning cycle.



- 9 Check the uECD output value on the control table. It should be lower than the first reading. If it is not, contact your Agilent service representative.
- 10 Reinstall the column.
- 11 Restore the analytical method.



11 Maintaining the NPD

Consumables and Parts for the NPD	206
Exploded Parts View of the NPD	209
Selecting an NPD jet	210
To Attach a Capillary Column Adapter on an Adaptable NPD	212
To Install a Capillary Column in the NPD	214
To Replace the NPD Bead Assembly	217
To Maintain the NPD Collector, Ceramic Insulators, and Jet	222
To Check the NPD Leakage Current	228



Consumables and Parts for the NPD

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Before selecting a jet, see “Selecting an NPD jet” on page 210.

Table 32 NPD parts

Description	Part number/quantity
Collector	G1534-20530
Screw, M3 × 0.5 × 8 mm	0515-0655
NPD white ceramic bead assembly	G1534-60570
NPD black ceramic bead assembly	5183-2007
Screw, M4 × 10 mm	0515-2495
J-clamp	1400-0015
NPD ceramic insulator kit <ul style="list-style-type: none"> • Metal O-rings, top and bottom • Ceramic insulators, upper and lower 	5182-9722
Insulation cup	19234-60720
NPD chemical sample kit solution of 0.65 ppm azobenzene, 1000 ppm octadecane, 1 ppm malathion in isooctane, 3 ampoules	18789-60060
NPD lid standoff	G1534-20590
Column adapters, for adaptable NPD only	
FID/NPD capillary column adapter	19244-80610
1/8-inch packed column adapter	19231-80520
1/4-inch packed column adapter	19231-80530
1/4-inch packed glass column adapter	G1532-20710
1/4-inch column nut	5180-4105 10/pk
1/4-inch Vespel/graphite ferrule	5080-8774 10/pk

Table 33 Jets for capillary optimized fittings

Jet type	Part number	Jet tip id	Length
Capillary with extended jet (recommended)	G1534-80580	0.29 mm (0.011 inch)	51.5 mm

Table 33 Jets for capillary optimized fittings (continued)

Jet type	Part number	Jet tip id	Length
Capillary	G1531-80560	0.29 mm (0.011 inch)	43 mm
High-temperature	G1531-80620	0.47 mm (0.018 inch)	43 mm

Table 34 Jets for adaptable fittings

Jet type	Part number	Jet tip id	Length
Capillary with extended jet (recommended)	G1534-80590	0.29 mm (0.11 inch)	70.5 mm
Capillary	19244-80560	0.29 mm (0.011 inch)	61.5 mm
Capillary, high-temperature	19244-80620	0.47 mm (0.018 inch)	61.5 mm
Packed	18710-20119	0.46 mm (0.018 inch)	63.6 mm

Table 35 Nuts, ferrules, and hardware for capillary columns


Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292

11 Maintaining the NPD

Table 35 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Exploded Parts View of the NPD




Selecting an NPD jet

Open the oven door and locate the column connection fitting at the base of the detector. It will look like either a capillary optimized fitting or an adaptable fitting.

Capillary optimized fitting



Adaptable fitting



- If you have an application that tends to clog the jet, select a jet with a wider tip id.
- When using packed columns in high column-bleed applications, the jet tends to clog with silicon dioxide.

For capillary optimized fittings, select one of the following from Table 36.

Table 36 Jets for capillary optimized fittings

Figure 3 ID	Jet type	Part number	Jet tip id	Length
1	Capillary with extended jet (recommended)	G1534-80580	0.29 mm (0.011 inch)	51.5 mm
2	Capillary	G1531-80560	0.29 mm (0.011 inch)	43 mm
3	High-temperature	G1531-80620	0.47 mm (0.018 inch)	43 mm



Figure 3 Capillary optimized NPD jets

For adaptable fittings, select one of the following from Table 37.

Table 37 Jets for adaptable fittings

Figure 4 ID	Jet type	Part number	Jet tip id	Length
1	Capillary with extended jet (recommended)	G1534-80590	0.29 mm (0.11 inch)	70.5 mm
2	Capillary	19244-80560	0.29 mm (0.011 inch)	61.5 mm
3	Capillary, high-temperature	19244-80620	0.47 mm (0.018 inch)	61.5 mm
4	Packed	18710-20119	0.46 mm (0.018 inch)	63.6 mm



Figure 4 Adaptable NPD jets

To Attach a Capillary Column Adapter on an Adaptable NPD

- 1 Gather the following materials:
 - Adapter (See “Consumables and Parts for the NPD” on page 206.)
 - 1/4-inch nut
 - 1/4-inch ferrule
 - Column cutter
 - 1/4-inch wrench
 - 9/16-inch open-end wrench
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.


WARNING

Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble a brass nut and Vespel/graphite ferrule onto the adapter.




- 4 Insert the adapter straight into the detector base as far as possible.

- 5 Hold the adapter in this position and finger-tighten the nut.

Adaptable fitting

Detector fitting

Adapter



- 6 Tighten an additional 1/4 turn with a wrench.

To Install a Capillary Column in the NPD

- 1 Gather the following materials :
 - Column
 - Ferrule(s) (See “Consumables and Parts for the NPD” on page 206.)
 - Column nut
 - Column cutter
 - 1/4-inch open-end wrench
 - Septum
 - Isopropanol
 - Lab tissue
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven, inlet, and/or detector may be hot enough to cause burns. If they are hot, wear heat-resistant gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

- 3 If using the adaptable detector, verify that the adapter is installed. (See “To Attach a Capillary Column Adapter on an Adaptable NPD” on page 212.)

Capillary optimized fitting




Adaptable fitting




CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 4 Place a septum, capillary column nut, and ferrule on the column.



- 5 Score the column using a glass scribing tool. The score must be square to ensure a clean break.



- 6 Break off the column end by supporting it against the column cutter opposite the scribe. Inspect the end with a magnifying loupe to make certain there are no burrs or jagged edges.




- 7 Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 8 Install the capillary column.

If the column id is greater than 0.1 mm:


- a Gently insert the column into the detector until it bottoms; do not attempt to force it further.
- b Finger-tighten the column nut, then withdraw the column about 1 mm. Tighten the nut an additional 1/4 turn with a wrench.

If the column id is 0.1 mm or less position the column so it extends above the ferrule by 48 mm (*capillary optimized fitting*) or 68 mm (*adaptable fitting*). Slide the septum up to hold the column nut and ferrule at this fixed position.

11 Maintaining the NPD



- c Insert the column into the detector. Slide the nut and ferrule up the column to the detector base. Finger-tighten the column nut until it grips the column.
- d Adjust the column (*not* the septum) position so that the septum is even with the bottom of the column nut. Tighten the nut an additional 1/4 turn with a wrench.



To Replace the NPD Bead Assembly

- 1 Gather the following:
 - Replacement NPD bead assembly. (See “Consumables and Parts for the NPD” on page 206.)
 - Lint-free gloves
 - T-10 Torx screwdriver

CAUTION

The ceramic bead is delicate. Be careful not to break or crack the bead. When performing maintenance on the NPD, avoid touching the bead with your fingers, and prevent it from coming in contact with other surfaces.

- 2 Set the NPD bead voltage to **0.0**. (Do not set it to **Off**.) Agilent data system users must use the keypad to set the voltage. To do this, the keyboard must be unlocked and you must close the GC parameters screen of the data system. After setting the voltage to **0.0**, save the data system method and shut down the instrument session.
- 3 Set **Adjust Offset** to **Off**.
- 4 Cool the detector to 60 °C or lower. Leave all gas flows on. To cool the detector faster, raise the GC detector cover and open the hinged NPD cover.
- 5 Remove the GC detector top cover.

WARNING

Hazardous voltages are present when the electronics top cover is open.

- 6 Remove the electronics top cover.
- 7 Put on lint-free gloves before touching any of the detector parts.


WARNING

Be careful! The oven or detector fittings may be hot enough to cause burns.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.


- 8** Twist the ring to disconnect the bead assembly cable. Push and twist the lock so that the button slides up in the groove, then pull the cable ends apart.




- 9** Remove the 3 T-10 Torx screws from the bead assembly.




- 10** Gently lift up and remove the old bead assembly. Avoid bumping the bead on the sides of the collector.




- 11** Remove the protective cap covering the new bead.
- 12** Mount the new bead assembly on the NPD lid. Be careful not to bump the bead on the sides of the lid or collector.
- 13** Replace the screws. Finger-tighten the first screw; tighten the remaining screws normally and then completely tighten the first screw. Do not overtighten the screws.



- 14** Carefully bend the bead assembly cable 90 °.



- 15** Reconnect the bead assembly cable to the NPD cable and twist the ring to lock the connection.



- 16** Close the NPD cover, install the GC detector top cover, and lower the electronics top cover. All covers must be closed to get a stable NPD baseline.
- 17** Restore normal NPD operating gas flows.
- 18** With all gases on, heat the detector to 150 °C and hold for about 15 minutes, then increase the temperature to 250 °C and hold for 15 minutes.
- 19** Increase the temperature to operating value (310 to 320 °C recommended). Allow 15 minutes for equilibration.
- 20** Check the NPD leakage current. (See “To Check the NPD Leakage Current” on page 228.) If > 2.0 pA, verify bead installation or see the 6890 Basic Troubleshooting manual.
- 21** If using an Agilent data system, connect to the instrument.
- 22** Restore the analytical method. Confirm the detector hydrogen, air, makeup gas flow rates.

- 23 Set equilibration time to **0.0**. Start the **Adjust offset** process. Enter the desired offset in the **Target offset** field. The default offset is 30 pA; an offset of 25 to 30 pA is sufficient for most applications. The bead life may be shortened at a higher offset.

To Maintain the NPD Collector, Ceramic Insulators, and Jet

When replacing the jet, always install a new collector, ceramic insulators, and metal C-rings.

When replacing the collector, Agilent recommends replacing the ceramic insulators and metal C-rings.

1 Gather the following:

- NPD ceramic insulator kit (See “Consumables and Parts for the NPD” on page 206.)
- Collector
- Cap for the bead
- T-10 and T-20 Torx screwdrivers
- Tweezers
- Cotton swab
- Solvent
- Methanol
- Jet (See “Selecting an NPD jet” on page 210.)
- Lint-free gloves
- Compressed, filtered dry air or nitrogen

CAUTION

The ceramic bead is delicate. Be careful not to break or crack the bead. When performing maintenance on the NPD, avoid touching the bead with your fingers, and prevent it from coming in contact with other surfaces.

- 2** Set the bead voltage to **0.0** and **Adjust Offset** to **Off**.
- 3** Check and note the NPD leakage current for reference. (See “To Check the NPD Leakage Current” on page 228.)
- 4** Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven or detector fittings may be hot enough to cause burns.

- 5** Remove the bead. (See “To Replace the NPD Bead Assembly” on page 217.)


CAUTION

This step exposes the interconnect spring. Be careful not to touch or disfigure the spring while working on the FID. Any dirt or bending will reduce the sensitivity of your detector.


CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 6 Remove the screws that secure the lid, and remove it. The top metal C-ring and upper ceramic insulator may be attached to the lid.



- 7 Remove the screws that secure the electrometer and the interconnect.



- 8 Pull the electrometer away from the detector to free the interconnect. Turn the electrometer to the right to obtain working space. Be careful not to touch or bend the spring. Be careful not to lose the EMI suppressor.
- 9 Remove the large metal O-ring and the upper ceramic insulator if they were not attached to the lid.

- 10** Remove the collector. If the detector is operated at high temperatures, the collector parts may stick inside the detector. Gently push and wiggle them to break the seal.




Figure 5 NPD collector, upper insulator, and metal O-ring

- 11** Use tweezers to remove the lower ceramic insulator and the two small metal O-rings located above and below the collector. If these parts are stuck together, do not separate them. If they are not stuck, remember which metal ring was on top of the insulator and which was below it. The pieces must be reassembled in the same orientation.





Figure 6 NPD lower ceramic insulator and metal O-rings

- 12** If not replacing the jet, skip to step 19.
13 Remove the column from the detector.
14 Loosen the jet with a nut driver.




- 15 Pull the jet straight out of the detector. Use tweezers, if necessary.

CAUTION


The adaptable NPD jet is longer than the capillary optimized NPD extended jet and should never be installed in a capillary optimized detector.

- 16 Place the jet in the detector body.
- 17 Using a nut driver, tighten the jet 1/6 turn past finger-tight. *Do not overtighten.*
- 18 Attach the column to the detector. (See “To Attach a Capillary Column Adapter on an Adaptable NPD” on page 212.)
- 19 Use a cotton swab wetted with solvent to clean the residue from the inside of the collector and around the jet. If the collector appears very dirty, replace it with a new one.




- 20 Install the bottom metal O-ring, the lower ceramic insulator, and the top metal O-ring. See Figure 5.
- 21 Install the collector.
- 22 Install the upper ceramic insulator and top metal C-ring above the collector. See Figure 6.


- 23** Install the lid, making sure that the NPD lid standoffs are in their slots. Hold the lid flat while each of the screws is tightened until they touch the lid. Tighten each screw evenly, 1/2 turn at a time, until tight. Do not overtighten.



- 24** Slide the electrometer interconnect into the slot on the lid and lower the electrometer into the mounting tray. Be careful not to touch or bend the spring.



- 25** Install the J-clamp and screws to secure the electrometer to the pallet.




- 26** Install the bead assembly and restore normal operating conditions. (See “To Replace the NPD Bead Assembly” on page 217.) (Do not reset the bead counter unless replaced.)

After installing new collector parts, the NPD leakage current should be lower. (See “To Check the NPD Leakage Current” on page 228.) If the leakage current is abnormal, check for proper reassembly of the detector (especially where the electrometer interconnect contacts the collector assembly) and for leaks.

To Check the NPD Leakage Current

- 1 Load the analytical method.
- 2 Set the **NPD Adjust Offset** to **Off** and the **Bead Voltage** to **0.00 V**.
 - Leave the NPD at operating temperature
 - Leave flows on or off
- 3 Press [**Front Detector**] or [**Back Detector**], then scroll to **Output**.
- 4 Verify that the output (leakage current) is stable and < 2.0 pA.

The output should slowly drop towards 0.0 pA, and should stabilize in the *tenths* of a picoamp. Current > 2.0pA indicates a problem.



12 Maintaining the FPD

Consumables and Parts for the FPD	230
Exploded Parts View of the FPD	232
To Install a Capillary Column Adapter in the FPD	233
To Attach a Capillary Column to the FPD	235
To Change the FPD Wavelength Filter	237
To Remove the FPD Vent Tube	240
To Replace the FPD Ignitor	242
To Install the FPD Vent Tube and Cover	244



Consumables and Parts for the FPD

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 38 FPD supplies

Description	Part number/quantity
Sulfur filter	1000-1437
Sulfur filter spacer	19256-20910
Phosphorus filter	19256-80010
Exit tube assembly, aluminum	19256-60700
Exit tube assembly, stainless steel	19256-20705
Vespel ferrule, 1/4-inch id	0100-1061
Ignitor replacement kit <ul style="list-style-type: none"> • O-ring • Spacer • Glow plug 	19256-60800
Screw, M3 × 66 mm, T-10	0515-0680
Collar	19256-20690
Capillary adapter nut	19256-21150
Capillary adapter seat	19256-21140
1/4-inch packed adapter	G1532-20710
Column measuring tool	19256-80640
Spring to secure photomultiplier tube	1460-1160
1/8-inch packed adapter nut	0100-0057
1/8-inch Vespel ferrule for packed adapter	0100-1332


Table 39 Nuts, ferrules, and hardware for capillary columns

Column id (mm)	Description	Typical use	Part number/quantity
.530	Ferrule, Vespel/graphite, 0.8-mm id	0.45-mm and 0.53-mm capillary columns	5062-3512 (10/pk)
	Ferrule, graphite, 1.0-mm id	0.53-mm capillary columns	5080-8773 (10/pk)
	Column nut, finger-tight (for 0.53-mm columns)	Connect column to inlet or detector	5020-8293

Table 39 Nuts, ferrules, and hardware for capillary columns (continued)

Column id (mm)	Description	Typical use	Part number/quantity
.320	Ferrule, Vespel/graphite, 0.5-mm id	0.32-mm capillary columns	5062-3514 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.250	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
.100 and .200	Ferrule, Vespel/graphite, 0.37-mm id	0.1-mm and 0.2-mm capillary columns	5062-3516 (10/pk)
	Ferrule, Vespel/graphite, 0.4-mm id	0.1-mm, 0.2-mm, and 0.25-mm capillary columns	5181-3323 (10/pk)
	Ferrule, graphite, 0.5-mm id	0.1-mm, 0.2-mm, 0.25-mm, and 0.32-mm capillary columns	5080-8853 (10/pk)
	Column nut, finger-tight (for .100- to .320-mm columns)	Connect column to inlet or detector	5020-8292
All	Ferrule, no-hole	Testing	5181-3308 (10/pk)
	Capillary column blanking nut	Testing—use with any ferrule	5020-8294
	Column nut, universal	Connect column to inlet or detector	5181-8830 (2/pk)
	Column cutter, ceramic wafer	Cutting capillary columns	5181-8836 (4/pk)

Exploded Parts View of the FPD



To Install a Capillary Column Adapter in the FPD

- 1 Gather the following:
 - FPD capillary column adapter (See “Consumables and Parts for the FPD” on page 230.)
 - Column cutter
 - 1/4-inch wrench
 - 9/16-inch wrench
 - Metric ruler
 - 1/8-inch nut
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Insert the capillary adapter into the 1/8-inch nut as shown, then thread the nut onto the detector fitting.

12 Maintaining the FPD



- 4 Finger-tighten the nut, then tighten an additional 1/8 turn with a wrench.

To Attach a Capillary Column to the FPD

- 1 Gather the following:
 - Column measuring tool (See “Consumables and Parts for the FPD” on page 230.)
 - Column cutter
 - 1/4-inch and 7/16-inch wrenches
 - Column nut
 - Ferrule
 - Capillary column
 - Lint-free gloves
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The oven and/or detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

WARNING


Wear safety glasses to protect your eyes from flying particles while handling, cutting, or installing glass or fused silica capillary columns. Use care in handling these columns to prevent puncture wounds.

CAUTION

Wear clean, lint-free gloves to prevent contamination of parts with dirt and skin oils.

- 3 Assemble a septum, column nut, and ferrule on the end of the column.
- 4 Insert the end of the column through the column measuring tool so that the end protrudes beyond the tool.

12 Maintaining the FPD



- 5 Tighten the column nut until it grips the column. Tighten the nut an additional $1/8$ to $1/4$ turn with a pair of wrenches. Snug the septum against the base of the column nut.
- 6 Use a wafer cutter at 45° to score the column.
- 7 Snap off the column end. The column may protrude about 1 mm beyond the end of the tool. Inspect the end with a magnifying loupe to make certain that there are no burrs or jagged edges.
- 8 Remove the column, nut, and swaged ferrule from the tool.
- 9 Wipe the column walls with a tissue dampened with isopropanol to remove fingerprints and dust.
- 10 Verify that a capillary adapter is installed in the detector fitting. (See "To Install a Capillary Column Adapter in the FPD" on page 233.)
- 11 Carefully thread the swaged column up into the adapter. Finger-tighten the column nut, then use a wrench to tighten an additional $1/8$ turn.

To Change the FPD Wavelength Filter

CAUTION

Do not touch the filter with your bare hands. For optimum performance and to avoid scratches, use lint-free gloves for assembling and inserting the filter into the assembly.

- 1 Gather the following:
 - Sulfur filter with filter spacer (See “Consumables and Parts for the FPD” on page 230.)
 - Phosphorus filter
 - Cotton swab
 - Lens tissue
 - Lint-free gloves
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

CAUTION


The photomultiplier tube (PMT) is extremely sensitive to light. Always turn off the electrometer (which turns off the high voltage to the PMT) before removing the PMT housing or opening the emissions chamber. Failing to do this can destroy the PMT.

Even with the electrometer off, protect the PMT from room light. Cap the housing after it is removed, place it end down to exclude light, or reduce the room light level before exposing the PMT. A brief exposure (always with the electrometer turned off) will not damage it, but prolonged exposure will cause a gradual loss of sensitivity.


- 3 Disconnect the retaining spring that holds the PMT assembly to the bracket. With a rotating motion, pull the assembly away from the filter housing.



- 4 To prevent light from damaging the PMT, cap the end or place it face down.



- 5 Place a clean cloth under the filter housing to catch the filter.
 - For phosphorus filter, use the sharpened wooden tip of a toothpick or cotton swab to dislodge the filter from the housing.
 - For sulfur filter (shown below), use the wooden tip of the cotton swab to remove the filter spacer. Then dislodge the filter from the housing.

**CAUTION**

Do not use cleaning fluids. Cleaning fluids will damage lens coatings.

- 6 Clean the new filter with lens tissue.

CAUTION

Filters are designed for the light of the flame to pass through in a specific direction. The triangle (on the edge of the phosphorus filter) and the arrow (on the edge of the sulfur filter) should face *away* from the flame and *toward* the PMT.

- 7 Install the filter in the filter housing. Install the sulfur filter spacer, if necessary.
- 8 Replace the PMT assembly and secure with the spring.
- 9 Restore the analytical method.

To Remove the FPD Vent Tube

- 1 Gather the following:
 - T-20 Torx screwdriver
 - 9/16-inch wrench
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

CAUTION

When turning off the GC, first turn off the flame to prevent condensation from dripping into the jet and column.



WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 3 Drain any water from the flexible vent tubing and discard.
- 4 Open the FPD cover.
- 5 Remove the flexible tubing from the FPD vent tube.
- 6 Loosen and remove the vent tube assembly with a wrench.



- 7 Remove the screws securing the FPD cover.
 - The single-wavelength detector has two screws at the bottom of the left side (top photo below) and two screws at the top of the right side (bottom photo below).
 - The dual-wavelength detector has two screws at the top of the right side (bottom photo below).



8 Lift the cover off the detector.


To Replace the FPD Ignitor

- 1 Gather the following:
 - Ignitor replacement kit. (See “Consumables and Parts for the FPD” on page 230.)
 - Torx screwdrivers, T-20 and T-10
 - 9/16-inch wrench
 - Tweezers
- 2 Prepare the detector for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

Be careful! The detector may be hot enough to cause burns. If the detector is hot, wear gloves to protect your hands.

- 3 Remove the vent tube assembly and cover. (See “To Remove the FPD Vent Tube” on page 240.)
- 4 Loosen the collar screw (some have two screws) holding the cable assembly to the ignitor. Remove the collar and cable assembly.




- 5 Use a wrench to loosen and remove the glow plug.

- 6 Remove the O-ring with tweezers.
- 7 Assemble the parts for the new ignitor.


Glow plug

Spacer

O-ring



- 8 Install the new ignitor assembly and tighten with a wrench. Do not overtighten.
- 9 Replace the ignitor collar and cable assembly and tighten the screw.



- 10 Replace the cover and the vent tube assembly. (See "To Install the FPD Vent Tube and Cover" on page 244.)
- 11 Restore the analytical method.
- 12 Wait 20 minutes for the detector to heat up, then ignite the flame.


To Install the FPD Vent Tube and Cover

- 1** Gather the following:
 - T-20 Torx screwdriver
 - 9/16-inch wrench
- 2** Install the cover.

Single-wavelength detector:

 - a** Start the two screws on the right side of the cover.
 - b** Start and tighten the screws at the base on the left side.
 - c** Tighten the screws on the right side.

Dual-wavelength detector: Install the cover (two screws).
- 3** Install the vent tube assembly.
- 4** Reconnect the flexible tubing to the vent tube assembly and route the open end to waste.
- 5** Close the FPD cover.



13 Maintaining a Valve

Consumables and Parts for Valves	246
Exploded Parts View of GC Rotary Valves	247
To Replace a Gas Sampling Valve Loop	248
To Align a Rotary Valve Rotor	250
To Replace a Rotary Valve in the Valve Box	251
To Remove the Upper Valve Box	254
To Install the Upper Valve Box	256




Consumables and Parts for Valves

See the Agilent catalog for consumables and supplies for a more complete listing, or visit the Agilent Web site for the latest information (www.agilent.com/chem/supplies).

Table 40 Valve supplies

Description	Part number
Valves, gas sampling	
6-port, 300 psi	0101-0584
6-port, 400 psi, 225 °C maximum temperature	5062-9508
6-port, 300 °C maximum temperature	0101-0460
6-port Hastelloy, 400 psi, 225 °C maximum temperature	5062-9509
10-port, 400 psi, 225 °C maximum temperature	5062-9510
10-port Nitronic 60, 300 psi, 350 °C maximum temperature	0101-0585
10-port Hastelloy, 400 psi, 225 °C maximum temperature	5062-9511
Valves, liquid sampling	
0.2- μ L, 1000 psi, stainless steel, 175 °C maximum temperature	0101-0636
0.5- μ L, 5000 psi, 175 °C maximum temperature	0101-0639
0.5- μ L, 1000 psi, stainless steel, 175 °C maximum temperature	0101-0637
1.0- μ L, 1000 psi, stainless steel, 175 °C maximum temperature	0101-0638
Gas sampling valve sample loops	
0.25-cc	0101-0303
0.50-cc	0101-0282
1.00-cc	0101-0299
2.00-cc	0101-0300
2.0-mL nickel loop, 1/16-inch	0101-0955
5.00-cc	0101-0301
10.00-cc	0101-0302
Ferrule, 1/16 inch stainless steel (10/pk)	5181-1291
Nut, 1/16 inch (10/pk)	5181-1292

Exploded Parts View of GC Rotary Valves



To Replace a Gas Sampling Valve Loop

- 1 Gather the following:
 - Replacement sample loop. (See “Consumables and Parts for Valves” on page 246.)
 - 1/4-inch wrench
 - Vacuum cleaner
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.
- 3 Turn off the detector.

WARNING

The oven, inlet, detector, and valve box may be very hot.

Sample and/or harmful gases may be present. Refer to your company's standard operating procedures for purging the chemicals from the sample line.


- 4 Set all valve box valves to **Off**.
- 5 Leave on the GC and valve actuator air.
- 6 Turn off the carrier gas and sample line flows and relieve any back pressure to the valve.

WARNING

The valve box insulation is made of refractory ceramic fibers (RCFs). To avoid inhaling RCF particles, we recommend these safety procedures:

1. Ventilate your work area
2. Wear long sleeves, gloves, safety glasses, and a disposable dust/mist respirator
3. Dispose of insulation in a sealed plastic bag
4. Vacuum any residual particles and discard
5. Wash your hands with mild soap and cold water after handling RCFs.

- 7 Remove the upper valve box. (See “To Remove the Upper Valve Box” on page 254.)
- 8 Vacuum any loose particulate insulation.
- 9 When the valve is cool, loosen the sample loop's two 1/4-inch fittings on the valve head and remove the loop.



- 10 Install the new sample loop.
- 11 Repressurize the sample loop and check for leaks.
- 12 Install the upper valve box. (See “To Install the Upper Valve Box” on page 256.)
- 13 Restore the analytical method.


To Align a Rotary Valve Rotor

- 1 Gather the following:
 - Flathead screwdriver
 - 3-mm hex key wrench
 - T-20 Torx screwdriver
- 2 Set the oven and valve box heated zones to a safe handling temperature (25 °C).
- 3 Set all valves to **Off**.

WARNING

The oven, inlet, detector, and valve box may be very hot. If they are hot, wear heat-resistant gloves to protect your hands.

- 4 Loosen the adjustment set screw.



- 5 Locate the rotor adjustment shaft on top of the actuator. Using a flathead screwdriver, rotate the valve rotor counterclockwise until it stops, then back it off a small amount to set one end of the rotor's motion (< 1 mm).
- 6 Tighten the adjustment set screw.
- 7 Turn the valve **On**, turn **Off** to check for smooth operation.
- 8 Restore the analytical method.

To Replace a Rotary Valve in the Valve Box

WARNING

Do not install a liquid sampling valve (LSV) in the valve box if you plan to heat the box above 75 °C. Heating an LSV over 75 °C can cause a leak and subsequent explosion. LSVs should be mounted in the side location to avoid potential explosions.

- 1 Gather the following:
 - Replacement valve (See “Consumables and Parts for Valves” on page 246.)
 - T-10 Torx screwdriver
 - 1/4-inch wrench
 - Needle-nosed pliers
 - Vacuum
- 2 Prepare the column and oven for maintenance. See “Preparing the GC for Maintenance” on page 15.

WARNING

The oven, inlet, detector, and valve box may be very hot.

Sample and/or harmful gases may be present. Refer to your company’s standard operating procedures for purging the chemicals from the sample line.


- 3 Set all valves to **Off**.
- 4 Leave on the GC and valve actuator air.
- 5 Turn off the carrier gas and sample line flows and relieve any back pressure to the valve.

WARNING


The valve box insulation is made of refractory ceramic fibers (RCFs). To avoid inhaling RCF particles, we recommend these safety procedures:

1. Ventilate your work area
2. Wear long sleeves, gloves, safety glasses, and a disposable dust/mist respirator
3. Dispose of insulation in a sealed plastic bag
4. Vacuum any residual particles and discard
5. Wash your hands with mild soap and cold water after handling RCFs.

- 6 Remove the upper valve box. (See “To Remove the Upper Valve Box” on page 254.) Vacuum any RCF insulation particulates from the valve box area.
- 7 Note the tubing connections to the existing valve and label if desired.
- 8 Disconnect the existing valve fittings.
- 9 Remove the two T-10 Torx screws attaching the valve to the valve box, then remove the valve from the valve box.
- 10 Place the new valve in the valve box. The gap in the index ring on top of a 6-port valve points toward the back of the GC if installed correctly. This is the **On** position. Install and tighten the two screws with a screwdriver.




- 11 Use needle-nosed pliers to move the valve rotor index pin of the valve counterclockwise until the pin touches the valve stop **Off** position.



- 12 Plumb the new valve using the existing fittings.

WARNING

Hazardous sample gases may be present.



- 13 Turn on the carrier and sample gases, then check for leaks at the valve fittings.
 - Using the needle-nosed pliers to toggle the valve, check both the **On** and **Off** positions.
 - When leak free, set the valve to **Off** (see step 11).
- 14 Install the upper valve box assembly. (See “To Install the Upper Valve Box” on page 256.)
- 15 Restore the analytical method.


To Remove the Upper Valve Box

- 1 Gather a T-20 Torx screwdriver.
- 2 Set the valve box to a safe handling temperature (25 °C).

WARNING

The oven, inlet, detector, and valve box may be very hot. If they are hot, wear heat-resistant gloves to protect your hands.

- 3 Lift and remove the detector cover.
- 4 Remove the mounting screws from the upper valve box.



- 5 Lift up and set aside.


WARNING

The valve box insulation is made of refractory ceramic fibers (RCFs). To avoid inhaling RCF particles, we recommend these safety procedures:


1. Ventilate your work area
 2. Wear long sleeves, gloves, safety glasses, and a disposable dust/mist respirator
 3. Dispose of insulation in a sealed plastic bag
 4. Vacuum any residual particles and discard
 5. Wash your hands with mild soap and cold water after handling RCFs.
-

To Install the Upper Valve Box

- 1 Gather the following:
 - T-20 Torx screwdriver
 - 3-mm hex key wrench
 - Flathead screwdriver
- 2 Verify that all valve rotors are in the full counterclockwise position (valve **Off**).
- 3 For each actuator that mates with a newly installed valve:
 - a Loosen the adjustment set screw.



- b Locate the rotor adjustment shaft on top of the actuator. Use a screwdriver to rotate the valve rotor counterclockwise until it stops.



- 4 Locate the two half-moon cutouts at the bottom back of the upper valve box. Place the upper valve box on top of the lower valve assembly, routing the heater/sensor wires through the cutouts. Secure with two T-20 mounting screws.
- 5 Push each coupling/shaft assembly downward with a flathead screwdriver until the slot on the coupling engages the rotor index pin.

If the coupling and valve do not engage, check that both are fully counterclockwise and try again. If necessary, turn the shaft slightly to engage the coupling.

- 6 For each newly installed valve:
 - a Using a flathead screwdriver, turn the rotor adjustment shaft counterclockwise until it stops, then back it off a small amount (< 1 mm) to set one end of the rotor's motion.
 - b Tighten the adjustment set screw.
- 7 Install the detector cover.
- 8 Restore normal operating condition.

13 Maintaining a Valve